

Review Form

Please complete a separate Review Form for each Docket.

Which Docket is being reviewed? (please underline)

Titles: NIOSH Manual of Analytical Methods (NMAM)

- 9106: *Methamphetamine and Illicit Drugs, Precursors, and Adulterants on Wipes by Liquid-Liquid Extraction (NIOSH-176)*
- 9109: *Methamphetamine and Illicit Drugs, Precursors, and Adulterants on Wipes by Solid Phase Extraction (NIOSH-177)*
- 0911: *Methamphetamine on Wipes by Liquid Chromatography-Mass Spectrometry-SIM (NIOSH-178)*.

Anticipated Publication: NIOSH Manual of Analytical Methods (NMAM), 5th Edition

Return by: September 30, 2009

Return to: Dr. W. Gregory Lotz, Director, Division of Applied Research and Technology, Mailstop R-2, NIOSH, 4676 Columbia Parkway, Cincinnati, OH 45226, or email at wlotz@cdc.gov.

YES NO
(explain below)

- | | | |
|--|-------------------------------------|-------------------------------------|
| 1. Does the Backup Data Report explain the problem and summarize relevant literature adequately? | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 2. Is the information in the Method and Backup Data Report technically accurate? | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 3. Are there any recommendations concerning organization of the documents? | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 4. Are there any changes or corrections needed in the Backup Data Report? | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 5. Are there any changes needed in the Method? | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 6. In general, is the Method and Backup Data Report satisfactory? | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 7. What is your recommendation for this Method as now written? (Check One): | | |
| a. Approve for publication/dissemination | <input type="checkbox"/> | <input type="checkbox"/> |
| b. Approve after modification (please describe) | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| c. Do not approve (please describe) | <input type="checkbox"/> | <input type="checkbox"/> |

DETAILED COMMENTS: (Provide comments in this space or on a separate sheet. Check here if a separate sheet is attached).

There is a table in both the Backup Data Report (Table IV) and method (Table II) that provides maximum surface contamination limits. The units in both are expressed as " μ /area". The correct units to express are " μ g/area."