


Frequently Asked Questions

For

NIOSH SAF (Standard Application Form) - Version 7

Question 1:

On the third form of the application, titled, “Application for the approval of respirators”, a “Type mismatch” error screen appears. What is the reason for this error and how do I correct it?



Answer:

This is due to a version compatibility problem. Notice the three buttons that are grayed out above are new in version 7 and not in version 6. The reason is that you are using the version 6 of the saf_data.mdb file, and it does not include the 3 fields introduced in version 7. That file was not removed from the directory before the installation of version 7. It has to be manually done, as the installation of version 7 does not replace the file.


Before installing version 7, the contents of your C:\NIOSH directory should be emptied, including all database (.mdb) and word files (.doc). If you already have version 6 or 7 installed, uninstall it, delete the contents of the C:\NIOSH directory and reinstall the program.

Question 2:

After completing the last screen in the application, I clicked on the “Done” button and file closed. Where can I access the application data?

Answer:

The application data is saved in the file saf_data.mdb that resides in the C:\NIOSH directory, which was specified in the installation. After completing the application, you will need to save the file as a different name. The most common name is something with the Applicant-Assigned Reference number (AAR) in it. This is done by going to the top menu and click File > Save As.



Attachment Checklist

Check the related documents which are included with this application:

<input type="checkbox"/> Quality Assurance Manual	<input type="checkbox"/> Exploded View Drawing
<input checked="" type="checkbox"/> Product Quality Control Plan	<input type="checkbox"/> Major Components Drawings
<input checked="" type="checkbox"/> Assembly Matrix	<input type="checkbox"/> Check
* <input type="checkbox"/> CGA Thread Specification	<input type="checkbox"/> User's Instructions
* <input checked="" type="checkbox"/> Burst Disc Pressure Range (SCBA only)	<input type="checkbox"/> Packaging Art
* <input checked="" type="checkbox"/> DOT Approval Documentation	<input type="checkbox"/> Special Gas Data
* <input checked="" type="checkbox"/> Lens meets GGG-M-125d requirement	<input type="checkbox"/> ESLI Test Data
<input checked="" type="checkbox"/> Service Life Plan	<input type="checkbox"/> List of Related Documents

All test data sufficient to demonstrate compliance with Part 84
 TC numbers are entered into the Reason for Application.

Draft Approval Labels:


<input type="checkbox"/> Air-Purifying Respirator	<input checked="" type="checkbox"/> Abbreviated Filter	<input type="checkbox"/> Scrubber Label
<input type="checkbox"/> Cartridge or Canister	<input checked="" type="checkbox"/> SCBA Label (in manual)	<input type="checkbox"/> SCBA Harness
<input type="checkbox"/> Filter	<input checked="" type="checkbox"/> SAR (in manual or on packaging)	
<input type="checkbox"/> Abbreviated Cartridge or Canister		

* Can be provided as a note on the relevant drawing.

Previous Done

I certify the information contained in this application is correct and that if approved, no further changes will be made to the product(s) without prior written approval of the National Institute for Occupational Safety and Health, Respirator Branch.

Initials of Authorized Rep.: _____



SAF Version 7

File Help

Open

Save As

Exit

NIOSH Standard Application Form, Version 7

National Institute for Occupational Safety and Health
National Personal Protective Technology Laboratory
Respirator Branch

NIOSH

STANDARD APPLICATION FOR THE CERTIFICATION OF RESPIRATORS

Applicant-Assigned Reference Number:

Type of Application:

New Approval

Extension of Approval

Quality Assurance Approval

Change "New" to "Extension"

Next View/Edit

Assign New# Print Preview

Reset Exit

Current Data File: c:\niosh\saf_data.mdb

OMB No. 0920-0109

Question 3:

Why does the screen below appear when I try to save or open my SAF program?



Answer:

The current version of SAF is not loaded on your computer. Download the latest version of SAF_v7 from the NIOSH SAF website.

<http://www.cdc.gov/niosh/npptl/resources/certpgmspt/standardapp.html>. This error occurs in previous versions of the SAF using the Microsoft run-time and the computer you are using is missing visual basic references, the most common being comdlg32.ocx. That version is no longer supported by NIOSH. You must have Microsoft Access to run SAF_v7. The following article explains one version of the error in detail.

<http://support.microsoft.com/kb/318597/en-us>

Question 4:

I am having problems when I try to uninstall a previous version of the SAF. I am being prompted for a file that is not on my computer. What do I do to solve this issue?

Answer:

Chances are that if you receive an error prompting you for a file during an uninstallation of SAF_v6 then the file is located on a CD mailed to you by NIOSH, in which you did the initial installation from. If you are prompted with the same screen during uninstallation of SAF_v7 then you either did the same installation from a CD or from a zip file within a compressed folder. Windows XP has built in .zip file handling and it may appear that you are in an uncompressed folder when you are not. The proper way to avoid this problem during an installation is to double click a zip file and copy and paste the contents to a new directory. If you do not have the medium or source of your installation and cannot uninstall your program, ask your system administrator to remove it from the registry manually.