


Note: All Dimensions in Millimeters

Figure 1-A (Revision 3) 3 April 2007 NIOSH CBRN Full Facepiece APR Mechanical Connector and Gasket

A External Thread



Note:

1. First and second thread must be properly formed in accordance with thread profile and tolerance.
2. No canister adapters permitted. A canister adapter is defined as any component between the 40mm internal mask threads and the 40mm external canister threads.