


# NSQAP



## 2018 ANNUAL SUMMARY REPORT

Newborn Screening  
Quality Assurance  
Program



Centers for Disease  
Control and Prevention  
National Center for  
Environmental Health

# Newborn Screening Quality Assurance Program **2018 Annual Summary Report, Volume 36**

.....  
U.S. Department of Health and Human Services  
Centers for Disease Control and Prevention  
National Center for Environmental Health  
**Division of Laboratory Sciences**



Note for accessibility: Explanations for Figure 2 and a general explanation for Figures 3–36 (bias plots) are located in [Appendix for Accessibility Descriptions, page 42](#).

## Contents

Introduction.....	1
About NSQAP.....	2
Countries Participating in NSQAP During 2018 .....	3
Filter Paper.....	5
Proficiency Testing.....	6
The Proficiency Testing Analytes .....	6
Proficiency Testing Materials and Methods.....	7
Proficiency Testing Data Handling.....	7
Proficiency Testing Errors.....	7
Non-Web Reported Analytes.....	11
Proficiency Testing Cutoff Values.....	14
Explanation of the NSQAP’s Grading Algorithm.....	23
2018 Bias Plots.....	24
Proficiency Testing Bias Plots.....	24
Appendix for Accessibility Descriptions.....	42
References.....	42
Acknowledgments.....	43

**NSQAP helps newborn screening laboratories ensure that testing accurately detects disorders, does not delay diagnoses, minimizes false-positive reports, and sustains high-quality performance.**



## Introduction

Newborn screening is one of the most successful preventative health programs in the United States. Healthcare professionals collect dried blood spot (DBS) specimens from more than 98% of all newborns shortly after birth in the United States. State and public health laboratories or their associated laboratories routinely screen these DBS specimens for certain genetic, metabolic, and endocrine disorders. The Centers for Disease Control and Prevention (CDC) Newborn Screening Quality Assurance Program (NSQAP) helps newborn screening laboratories with these testing processes.

NSQAP produces certified DBS materials for proficiency testing (PT) and quality control (QC) analysis, works to improve the quality and scope of laboratory services, and provides consultation to laboratories. State-operated and private newborn screening laboratories process thousands of DBS specimens daily. NSQAP helps newborn screening laboratories ensure that testing accurately detects disorders, does not delay

diagnoses, minimizes false-positive reports, and sustains high-quality performance.

CDC's Newborn Screening and Molecular Biology Branch (NSMBB) has been granted International Organization for Standardization (ISO)/International Electrotechnical Commission (IEC) 17043 accreditation by the American Association for Laboratory Accreditation (A2LA). Accreditation was achieved after a thorough review of its quality management system and competence to develop and administer specific PT protocols. The branch's NSQAP web-based PT programs are included in the A2LA Scope of Accreditation. The scope of accreditation does not include testing for glucose-6-phosphate dehydrogenase (G6PD) and NSQAP non-web-based PT programs. Consult A2LA Certificate#4190.01 for a list of accredited NSMBB PT programs.



## About NSQAP


For more than 40 years, NSQAP and its cosponsor the Association of Public Health Laboratories, have researched the development of DBS quality assurance materials for newborn screening tests and have assisted laboratories with DBS-related testing issues. NSQAP primarily supports U.S. newborn screening laboratories. Private and international laboratories may also enroll in the program. Participation is voluntary. NSQAP provides quality assurance services for the core (primary) and secondary conditions listed in the U.S. Recommended Uniform Screening Panel (RUSP) [1].

Over the years, NSQAP services and participation have grown substantially. In 2018, active program participants included 663 newborn screening laboratories in 85 countries (at least one laboratory per country) (Figure 1). Of these laboratories, 588 participated in PT (Table 1) and 522 in QC (Table 2). The program distributed DBS materials for 78 analytes to participating laboratories (Tables 1 and 2).


To offer more specialized services, NSQAP works with the Biochemical Mass Spectrometry Laboratory (BMSL) and the Molecular Quality Improvement Program (MQIP) in the Newborn Screening and Molecular Biology Branch.


BMSL offers newborn screening tandem mass spectrometry (MS/MS) services, education, and research opportunities. It also oversees the amino acids, acylcarnitines, biotinidase, total galactose (TGal), galactose-1-phosphate uridylyltransferase (GALT), G6PD, Lysosomal Storage Disorders (LSD), and the filter paper evaluation programs.

MQIP oversees the Cystic Fibrosis DNA (CFDNA) and T-cell Receptor Excision Circle (TREC) PT programs and assists newborn screening laboratories with molecular testing. It also offers the Molecular Assessment Program (MAP) which conducts site visits to U.S. newborn screening laboratories that carry out molecular testing. These visits assess components of molecular testing and include program-tailored guidance for laboratory-specific needs and assistance in evaluating ongoing and future molecular testing procedures.



**Figure 1.** Eighty-five countries participated in the Newborn Screening Quality Assurance Program in 2018.

 Countries Shown on World Map that Participated in NSQAP During 2018



- |            |                |            |             |                 |                      |
|------------|----------------|------------|-------------|-----------------|----------------------|
| Argentina  | Cuba           | India      | Malaysia    | Poland          | Tanzania             |
| Armenia    | Czech Republic | Indonesia  | Malta       | Portugal        | Thailand             |
| Australia  | Denmark        | Iraq       | Mexico      | Qatar           | Tunisia              |
| Austria    | Ecuador        | Ireland    | Mongolia    | Romania         | Turkey               |
| Bahrain    | Egypt          | Israel     | Morocco     | Saudi Arabia    | Ukraine              |
| Belgium    | El Salvador    | Italy      | Netherlands | Singapore       | United Arab Emirates |
| Bolivia    | Estonia        | Japan      | New Zealand | Slovak Republic | United Kingdom       |
| Brazil     | Finland        | Jordan     | Nigeria     | Slovenia        | United States        |
| Bulgaria   | France         | Kazakhstan | Norway      | South Africa    | Uruguay              |
| Canada     | Germany        | Kuwait     | Oman        | South Korea     | Vietnam              |
| Chile      | Greece         | Latvia     | Pakistan    | Spain           |                      |
| China      | Guatemala      | Lebanon    | Panama      | Sri Lanka       |                      |
| Colombia   | Honduras       | Lithuania  | Paraguay    | Sweden          |                      |
| Costa Rica | Hungary        | Luxembourg | Peru        | Switzerland     |                      |
| Croatia    | Iceland        | Macedonia  | Philippines | Taiwan          |                      |





**Table 1.** Number of participants reporting proficiency testing analytes, 2018 (N = 588)

Analyte	Total PT Participation in 2018	Analyte	Total PT Participation in 2018
170HP	303	C6	332
T4	93	C8	358
TSH	377	C10	345
TGal	193	C10:1	313
BIOT	219	C10:2	224
GALT	148	C14	330
IRT	252	C14:1	338
G6PD	101	C16	340
CFDNA	71	C160H	337
Hb	70	C18	323
Anti-HIV-1	22	C18:1	311
TOXO	18	C180H	286
TREC	62	170HP2	24
Arg	304	4AD2	24
Cit	327	CORT2	24
Leu	358	11D2	17
Met	342	21D2	17
Phe	461	GALC	11
SUAC	163	GAA	17
Tyr	359	IDUA	15
Val	321	C24-LPC	17
C0(L)	347	C26-LPC	20
C3	344		
C3DC	145		
C3DC+C40H	143		
C4	324		
C40H	138		
C5	357		
C5:1	317		
C5DC	340		
C50H	310		

**Table 2.** Number of participants reporting quality control analytes, 2018 (N = 522)

Analyte	Total QC participation in 2018	Analyte	Total QC participation in 2018
170HP	264	C160H	303
T4	77	C18	302
TSH	340	C180H	258
TGal	166	170HP2	23
GALT	93	4AD2	23
IRT	210	CORT2	23
Ala	260	11D2	16
Arg	285	21D2	16
Cit	305	GALC	19
Gly	227	GAA	31
Leu	319	IDUA	28
Met	313	GLA	27
Orn	231	ABG	26
Phe	389	ASM	14
SUAC	142	C20-LPC	7
Tyr	320	C22-LPC	7
Val	302	C24-LPC	15
C0	308	C26-LPC	19
C2	303	GAA2	8
C3	308	CRE2	5
C3DC	133	ALE2	20
C3DC+C40H	151	ILE2	18
C4	302	LEU2	19
C40H	127	PHE2	22
C5	316	TYR2	21
C5DC	304	VAL2	21
C50H	280	MMA2	22
C6	307	EMA2	9
C8	317	MCA2	16
C10	313	MA2	2
C12	297	tHCY2	21
C14	309		
C16	310		



## Filter Paper

NSQAP evaluates absorption characteristics of all filter paper lots approved by the Food and Drug Administration (FDA) as a newborn screening collection device [3]. Filter paper manufacturers must establish their own parallel evaluation. NSQAP's evaluations are an impartial and voluntary service offered as a function of our QC program; they do not constitute endorsement of any product.

The disk punched from a DBS specimen gives a volumetric measurement that requires a high degree of uniformity among and within production lots. NSQAP uses an isotopic method developed at CDC to evaluate and compare filter paper lots. It equates mean counts per minute of added radioisotope-labeled thyroxine (T4) contained within a 3.2-mm disk with the serum absorption volume of the disks made from washed, intact red blood cells (RBCs). The latest version of Clinical Laboratory Standards Institute (CLSI) Standard NBS01-A6, Blood Collection on Filter Paper for Newborn Screening Programs, describes the method.

FDA-approved newborn screening filter paper manufacturers (GE Healthcare Biosciences Corporation and PerkinElmer Health Sciences) provide NSQAP with statistically valid sample sets of unprinted filter paper from each production lot. Tables 3 and 4 show serum absorption volumes from 10 most recent lots of these two filter paper sources. The published standardized acceptable serum absorption volume per 3.2-mm disk (mean value and 95% confidence interval) is  $1.44 \pm 0.20$   $\mu\text{L}$  of washed intact RBCs [3]. The testing results in Tables 3 and 4 are informational only. Each mean value is within the acceptable range for the matrix used. All lots are homogenous (i.e., the measured within-spot, within-sheet, and among-sheets variances were within acceptable limits). CDC used 903™ filter paper lots W152, W161, and W171 to produce the QC and PT specimens distributed in 2018.

**Table 3.** PerkinElmer 226 specimen collection filter paper absorption characteristics by lot number—intact red cells

<b>Filter Paper</b>	<b>Date of Evaluation</b>	<b>Serum Volume (<math>\mu\text{L}</math>) per 3.2 mm (1/8") Punch</b>	<b>Absorption Time (sec)</b>	<b>Spot Diameter (mm)</b>
<b>Lot No.</b>	<b>Month/Year</b>	<b>Average (StDev)</b>	<b>Average (StDev)</b>	<b>Average (StDev)</b>
<b>112147</b>	Sept 2018	1.49 (0.11)	7.9 (0.9)	15.8 (0.6)
<b>111064</b>	July 2017	1.47 (0.20)	8.2 (1.0)	15.7 (0.5)
<b>110092</b>	July 2016	1.45 (0.09)	9.0 (1.2)	16.0 (0.7)
<b>105617</b>	May 2016	1.46 (0.08)	8.3 (1.8)	15.8 (0.5)
<b>105616</b>	Jan 2016	1.56 (0.11)	10.6 (2.0)	15.6 (0.5)
<b>105178</b>	Aug 2015	1.46 (0.09)	7.8 (1.1)	15.9 (0.6)
<b>104568</b>	March 2015	1.56 (0.10)	10.1 (2.1)	15.9 (0.7)
<b>103649</b>	March 2014	1.53 (0.10)	9.7 (3.1)	15.7 (0.7)
<b>102928</b>	Aug 2013	1.38 (0.09)	8.5 (0.9)	16.1 (0.5)
<b>102277</b>	Dec 2012	1.47 (0.11)	13.0 (4.9)	15.8 (0.6)



**Table 4.** 903™ specimen collection filter paper absorption characteristics by lot number—intact red cells

Filter Paper	Date of Evaluation	Serum Volume (µL) per 3.2 mm (1/8") Punch	Absorption Time (sec)	Spot Diameter (mm)
Lot No.	Month/Year	Average (StDev)	Average (StDev)	Average (StDev)
<b>W181</b>	Sept 2018	1.42 (0.12)	16.1 (3.3)	16.2 (0.6)
<b>W171</b>	April 2017	1.39 (0.10)	19.7 (4.7)	16.0 (0.7)
<b>W162</b>	Jan 2017	1.43 (0.08)	12.9 (2.7)	16.0 (0.7)
<b>W161</b>	May 2016	1.41 (0.08)	14.8 (3.7)	16.2 (0.8)
<b>W152</b>	Aug 2015	1.37 (0.09)	15.8 (2.4)	16.2 (0.6)
<b>W151</b>	Aug 2015	1.39 (0.08)	15.2 (2.6)	16.2 (0.8)
<b>W142</b>	April 2015	1.46 (0.08)	11.0 (2.2)	16.0 (0.7)
<b>W141</b>	March 2014	1.53 (0.10)	13.8 (3.6)	15.9 (0.6)
<b>W131</b>	Aug 2013	1.40 (0.07)	10.4 (1.4)	16.1 (0.5)
<b>W122</b>	May 2013	1.41 (0.11)	14.8 (2.9)	16.3 (0.5)

## Proficiency Testing

NSQAP distributes PT materials at least three times per year. PT panels consist of five blind-coded 75µL DBS specimens. Specimen sets are packaged in a zip-closed, metalized plastic bag with desiccant. Instructions for

analysis and reporting data are located online at [https://www.cdc.gov/labstandards/nsgap\\_resources.html](https://www.cdc.gov/labstandards/nsgap_resources.html). These specimens provide an independent, external assessment of each laboratory's performance.

### The Proficiency Testing Analytes

#### AMINO ACIDS

- arginine (Arg)
- citrulline (Cit)
- leucine (Leu)
- methionine (Met)
- phenylalanine (Phe)
- succinylacetone (SUAC)
- tyrosine (Tyr)
- valine (Val)

#### ACYLCARNITINES

- low free carnitine (C0(L))
- propionylcarnitine (C3)
- malonylcarnitine (C3DC)
- butyrylcarnitine (C4)
- hydroxybutyrylcarnitine (C4OH)
- isovalerylcarnitine (C5)
- tiglylcarnitine (C5:1)

- glutarylcarnitine (C5DC)
- hydroxyisovalerylcarnitine (C5OH)
- hexanoylcarnitine (C6)
- octanoylcarnitine (C8)
- decanoylcarnitine (C10)
- decenoylcarnitine (C10:1)
- decadienoylcarnitine (C10:2)
- myristoylcarnitine (C14)
- tetradecenoylcarnitine (C14:1)
- palmitoylcarnitine (C16)
- hydroxypalmitoylcarnitine (C16OH)
- stearoylcarnitine (C18)
- oleoylcarnitine (C18:1)
- Hydroxystearoylcarnitine (C18OH)

#### OTHER ANALYTES

- 17 α-hydroxyprogesterone (17OHP)
- 24:0-lysophosphatidylcholine (C24-LPC)
- 26:0-lysophosphatidylcholine (C26-LPC)
- anti-HIV-1 Antibodies (HIV)
- acid-α-glucosidase (GAA)
- α-L-iduronidase (IDUA)
- biotinidase (BIOT)
- cystic fibrosis DNA (CFDNA)
- Galactose-1-phosphate Uridyltransferase (GALT)
- galactocerebrosidase (GALC)
- glucose-6-phosphate dehydrogenase (G6PD)
- immunoreactive trypsinogen (IRT)
- Total Galactose (TGal)
- second-tier 17 α-hydroxyprogesterone (17OHP2)
- second-tier 4-androstenedione (4AD2)
- second-tier cortisol (CORT2)
- second-tier 11-deoxycortisol (11D2)
- second-tier 21-deoxycortisol (21D2)
- sickle cell and other hemoglobinopathies (Hb)
- T-cell receptor excision circle (TREC)
- Thyroid Stimulating Hormone (TSH)
- thyroxine (T4)
- anti-*Toxoplasma* Antibodies (TOXO)

## Proficiency Testing Materials and Methods

NSQAP certifies PT specimens for homogeneity, accuracy, stability, and suitability for newborn screening assays. Most PT specimens are prepared from whole blood of 50% hematocrit. PT materials are produced from one of the following: using unaltered donor blood, enriching a single blood unit or pooling blood units.

**Purified analytes** are used for PT enrichments. Enrichments made with commercially available or custom-synthesized analytes are based on weight. Small variances in enrichments and recoveries might result from impurities in the purchased (synthesized) materials and endogenous analyte concentrations.

**Congenital hypothyroid PT specimens** are enriched with measured amounts of T4 and TSH after reconstituting washed RBCs with purchased T4-depleted charcoal-stripped serum.

**IRT PT specimens** are made from a washed, hematocrit-adjusted blood that is treated with a protease inhibitor then enriched with commercially-purchased IRT.

**TGal materials** are enriched with galactose and galactose-1-phosphate, allowing measurement of free galactose (galactose alone) and total galactose (free galactose plus galactose-1-phosphate).

**Biotinidase PT** specimens are made using heat-treated serum combined with compatible donor RBCs.

**Deficient GALT PT** specimens are made using a 50/50 saline/serum solution combined with compatible washed RBCs and then heat-treating the pool.

**Low free carnitine (CO[L])** materials are produced by washing fresh RBCs at least six times then combining with charcoal-stripped serum.

**CFDNA PT** specimens are prepared using blood from anonymous cystic fibrosis patients, carriers, or unaffected individuals without hematocrit adjustment.

**Hemoglobin** specimens are made from hematocrit-adjusted individual umbilical cord blood units.

**Anti-HIV-1 antibody PT** specimens are prepared by mixing purchased donor serum reactive for HIV-1 antibodies and washed RBCs to achieve the desired reactivity.

**T-cell receptor excision circle PT** specimens are prepared from human blood, including cord blood from

unaffected persons and modified adult blood depleted of mononuclear cells or leukocytes.

**Lysosomal storage disorder** specimens are prepared from human blood, including cord blood from unaffected persons and leukodepleted adult blood restored with lymphoblast cell lines derived from patients with LSD.

**Toxoplasma immunoglobulin G and M DBS** specimens are prepared by combining human serum samples collected from patients exposed to *Toxoplasma gondii* with compatible washed RBCs.

## Proficiency Testing Data Handling

Participants submit PT data and clinical assessment through the NSQAP data reporting website or use an Excel data reporting form downloaded from the NSQAP section of the CDC website at [https://www.cdc.gov/labstandards/nsqap\\_resources.html](https://www.cdc.gov/labstandards/nsqap_resources.html).

Laboratories that submit results before the data reporting deadline will receive an individual laboratory evaluation and their data are included in the data summary report.

## Proficiency Testing Errors

Screening programs are designed to minimize false-negative reports, but this precautionary approach could result in false-positive misclassifications. Laboratories should monitor false-positive misclassifications to keep them as low as possible.

Tables 5–7 show the PT errors reported in 2018 by domestic and international laboratories for qualitative assessments by disorder/analyte. Because of specific clinical assessment practices, presumptive clinical classifications (qualitative assessments) of some specimens might differ by participant. If participants provided their cutoff values, those values were applied in the final evaluation of the error judgment (Figure 2). The rates for false-positive misclassifications were based on the number of negative specimens tested. The rates for false-negative misclassifications were based on the number of positive specimens tested.

The results of some PT specimens were near the decision level for clinical assessment. This rigorously tested the ability of laboratories to make the expected cutoff decision. Most specimens near the mean cutoff value are classified as not-evaluated specimens. As such, they were not included in the error calculations.


**Table 5.** Summary of non-MS/MS proficiency test errors by domestic and international laboratories

**Domestic**

Analyte	Positive specimens assayed (N)	False negative errors (%)	Negative specimens assayed (N)	False positive errors (%)
<b>Congenital adrenal hyperplasia</b>	126	0.0%	504	0.0%
<b>Biotinidase deficiency</b>	215	1.4%	430	1.9%
<b>G6PD deficiency</b>	15	0.0%	30	3.3%
<b>GALT deficiency</b>	258	0.0%	387	0.0%
<b>Immunoreactive trypsinogen</b>	262	1.1%	393	0.0%
<b>Congenital hypothyroidism</b>	215	0.0%	430	0.2%
<b>Galactosemia</b>	92	0.0%	253	0.0%

**International**

Analyte	Positive specimens assayed (N)	False negative errors (%)	Negative specimens assayed (N)	False positive errors (%)
<b>Congenital adrenal hyperplasia</b>	658	1.7%	2637	0.7%
<b>Biotinidase deficiency</b>	752	1.7%	1493	2.0%
<b>G6PD deficiency</b>	404	2.5%	836	1.2%
<b>GALT deficiency</b>	548	2.4%	822	1.1%
<b>Immunoreactive trypsinogen</b>	1038	1.3%	1557	6.3%
<b>Congenital hypothyroidism</b>	1408	0.7%	2797	0.6%
<b>Galactosemia</b>	560	2.0%	1550	0.3%



**Table 6.** Summary of amino acid and acylcarnitine proficiency test errors by **Domestic** laboratories

Analyte	Positive specimens assayed (N)	False negative errors (%)	Negative specimens assayed (N)	False positive errors (%)
<b>Arginine screen</b>	36	0.0%	504	0.0%
<b>Citrulline screen</b>	132	0.0%	528	0.0%
<b>Leucine screen</b>	89	0.0%	581	0.0%
<b>Methionine screen</b>	176	0.0%	484	0.0%
<b>Phenylalanine screen</b>	164	1.2%	656	1.2%
<b>Succinylacetone screen</b>	34	0.0%	486	0.0%
<b>Tyrosine screen</b>	100	0.0%	635	0.3%
<b>Valine screen</b>	90	3.3%	365	1.4%
<b>C0(L) screen</b>	0	0.0%	700	0.4%
<b>C3 screen</b>	96	1.0%	619	0.0%
<b>C3DC screen</b>	0	0.0%	280	0.0%
<b>C3DC+C40H screen</b>	0	0.0%	320	0.0%
<b>C4 screen</b>	0	0.0%	645	0.5%
<b>C40H Screen</b>	0	0.0%	270	0.0%
<b>C5 screen</b>	48	0.0%	667	0.0%
<b>C5:1 screen</b>	0	0.0%	700	0.1%
<b>C5DC screen</b>	93	1.1%	607	0.0%
<b>C50H screen</b>	47	2.1%	653	0.3%
<b>C6 screen</b>	134	0.7%	536	0.0%
<b>C8 screen</b>	96	0.0%	619	0.2%
<b>C10 screen</b>	88	4.5%	567	0.0%
<b>C10:1 screen</b>	83	7.2%	532	0.0%
<b>C10:2 screen</b>	0	0.0%	405	0.0%
<b>C14 screen</b>	130	5.4%	525	0.0%
<b>C14:1 screen</b>	140	0.0%	570	0.0%
<b>C16 screen</b>	0	0.0%	675	0.0%
<b>C160H screen</b>	47	0.0%	668	0.0%
<b>C18 screen</b>	0	0.0%	630	0.0%
<b>C18:1 screen</b>	0	0.0%	610	0.0%
<b>C180H screen</b>	0	0.0%	545	0.2%



**Table 7.** Summary of amino acid and acylcarnitine proficiency test errors by **International** laboratories

<b>Analyte</b>	<b>Positive specimens assayed (N)</b>	<b>False negative errors (%)</b>	<b>Negative specimens assayed (N)</b>	<b>False positive errors (%)</b>
<b>Arginine screen</b>	219	2.3%	3196	0.3%
<b>Citrulline screen</b>	733	1.9%	2932	0.6%
<b>Leucine screen</b>	554	2.7%	3486	0.4%
<b>Methionine screen</b>	1034	1.8%	2826	0.8%
<b>Phenylalanine screen</b>	1057	2.4%	4228	1.4%
<b>Succinylacetone screen</b>	91	0.0%	1424	0.4%
<b>Tyrosine screen</b>	508	0.8%	3492	1.9%
<b>Valine screen</b>	772	4.8%	2968	1.3%
<b>C0(L) screen</b>	0	0.0%	3950	1.4%
<b>C3 screen</b>	515	1.7%	3380	0.4%
<b>C3DC screen</b>	0	0.0%	1590	2.5%
<b>C3DC+C40H screen</b>	0	0.0%	1465	0.5%
<b>C4 screen</b>	0	0.0%	3645	1.3%
<b>C40H screen</b>	0	0.0%	1490	0.8%
<b>C5 screen</b>	281	6.4%	3784	0.1%
<b>C5:1 screen</b>	0	0.0%	3485	1.7%
<b>C5DC screen</b>	503	0.4%	3332	1.1%
<b>C50H screen</b>	232	6.0%	3173	2.6%
<b>C6 screen</b>	750	0.9%	3000	0.8%
<b>C8 screen</b>	543	0.7%	3552	1.3%
<b>C10 screen</b>	526	4.2%	3439	0.8%
<b>C10:1 screen</b>	461	8.0%	3029	0.8%
<b>C10:2 screen</b>	0	0.0%	2455	0.6%
<b>C14 screen</b>	741	5.7%	2979	0.6%
<b>C14:1 screen</b>	759	1.3%	3041	0.4%
<b>C16 screen</b>	0	0.0%	3825	0.8%
<b>C160H screen</b>	257	1.6%	3528	0.8%
<b>C18 screen</b>	0	0.0%	3640	0.3%
<b>C18:1 screen</b>	0	0.0%	3480	0.3%
<b>C180H screen</b>	0	0.0%	3150	1.3%

## Non-Web Reported Analytes

Table 8 shows a summary of PT errors for programs not reported on the NSQAP database website. Those include the Hb, CFDNA Variant Detection, LSD, TREC, Anti-*Toxoplasma* Antibodies, X-linked Adrenoleukodystrophy (XALD), and Second-tier Congenital Adrenal Hyperplasia (CAH) programs.

The CFDNA PT program provides evaluations based on allele identification and clinical assessment. Allele

detection is dependent on the method used. Table 9 summarizes the CF variant challenges distributed in 2018.

Table 10 shows the challenges distributed in 2018 for sickle cell disease and other hemoglobinopathies. Participants are evaluated on hemoglobin phenotypes and ability to provide correct clinical assessments.

**Table 8.** Summary of non-web based analyte proficiency test errors

### Sickle Cell and Other Hemoglobinopathies

Proficiency Test	Domestic	International
<b>Specimens Assayed</b>	665	370
<b>Phenotype Errors</b>	0.2%	2.7%
<b>Clinical Assessment Errors</b>	0.0%	2.4%

### Cystic Fibrosis DNA Variant

Proficiency Test	Domestic	International
<b>Specimens Assayed</b>	638	700
<b>Allele Errors</b>	0.5%	1.7%
<b>Clinical Assessment Errors</b>	0.2%	0.6%

### Lysosomal Storage Disorders

#### Krabbe

Proficiency Test	Domestic	International
<b>Specimens Assayed</b>	155	n/a
<b>Clinical Assessment Errors</b>	1.3%	n/a

#### Pompe

Proficiency Test	Domestic	International
<b>Specimens Assayed</b>	235	n/a
<b>Clinical Assessment Errors</b>	0.9%	n/a

#### Mucopolysaccharidosis Type I

Proficiency Test	Domestic	International
<b>Specimens Assayed</b>	205	n/a
<b>Clinical Assessment Errors</b>	0.0%	n/a

## T-cell Receptor Excision Circle

Proficiency Test	Domestic	International
Total Specimens Assayed	1466	580
Clinical Assessment Errors	0.7%	3.3%

---

## Second-tier Congenital Adrenal Hyperplasia

Proficiency Test	Domestic	International
Specimens Assayed	75	300
Clinical Assessment Errors	5.3%	9.7%

---

## X-linked Adrenoleukodystrophy

### 24:0 Lysophosphatidylcholine

Proficiency Test	Domestic	International
Specimens Assayed	165	60
Clinical Assessment Errors	1.2%	0.0%

---

### 26:0 Lysophosphatidylcholine

Proficiency Test	Domestic	International
Specimens Assayed	165	75
Clinical Assessment Errors	1.2%	0.0%

---

**Table 9.** Cystic Fibrosis DNA variant (CTFR gene) challenges distributed in 2018

Variant (Legacy name)	Variant (HGVS nomenclature)	Variants sent
<b>F508del</b>	(c.1521_1523delCTT)	10
<b>W1204X</b>	(c.3612G>A)	2
<b>R1066C</b>	(c.3196C>T)	1
<b>G551D</b>	(c.1652G>A)	1
<b>2789+5G&gt;A</b>	(c.2657+5G>A)	1
<b>3876delA</b>	(c.3744delA)	1
<b>R334W</b>	(c.1000C>T)	1
<b>Y1092X</b>	(c.3276C>A)	1
<b>1898+G&gt;A</b>	(c.1766+1G>A)	1
<b>3120+1G&gt;A</b>	(c.2988+1G>A)	1
<b>Q890X</b>	(c.2668C>T)	1
<b>E60X</b>	(c.178G>T)	1
<b>G542X</b>	(c.1624G>T)	1
<b>3849+10kbC&gt;T</b>	(c.3717+12191C>T)	1
<b>2183AA&gt;G</b>	(c.2051_2052delAAinsG)	1

**Table 10.** Hemoglobinopathies accepted presumptive phenotype distribution

Quarter	Specimen 1	Specimen 2	Specimen 3	Specimen 4	Specimen 5
<b>Quarter 1</b>	FAS	FA	FAC	Alpha Thal— Silent Carrier, FA	FA
<b>Quarter 3</b>	FAS	FAC	FA	FS, FSU	FA
<b>Quarter 4</b>	FS	FA	FAC	FAS	FA



## Proficiency Testing Cutoff Values

Participants report the decision level for sorting test results as presumptive positive (outside normal limits) from results reported as negative (within normal limits), based on their established cutoff value. CDC does not test newborns; therefore, establishing a population cutoff value is not possible. Therefore, CDC cutoff values are determined by using the mean of all domestic laboratory cutoff values. (Note: Each laboratory should establish its own cutoff values rather than using the CDC reported cutoff values.)

For PT evaluations, the participating laboratory's reported cutoff value is applied to our grading

algorithm. If no cutoff value is reported for a particular analytical result, the grading algorithm will default to the NSQAP-assigned cutoff value, which is based on the domestic mean cutoff value. (Figure 2)

Tables 11–13 summarize the reported cutoff values for domestic and international laboratories. The tables show the values for mean, median, and mode for each analyte. Tables 14–16 summarize the mean, median, mode, and minimum/maximum for reported domestic cutoff values, by method.

**Table 11.** Summary of non-MS/MS cutoff values for domestic and international laboratories, 2018

### Domestic

Analyte	N	Mean	Median	Mode	Minimum	Maximum
<b>17OHP (ng/mL serum)</b>	41	33.7	33.0	30.0	17.8	65.0
<b>IRT (ng/mL blood)</b>	42	68.3	63.4	60.0	46.3	160.0
<b>T4 (µg/dL serum)</b>	21	6.4	6.0	5.0	5.0	8.5
<b>TGal (mg/dL blood)</b>	23	11.2	10.0	10.0	6.0	20.0
<b>TSH (µIU/mL serum)</b>	42	30.7	25.0	20.0	19.0	58.0
<b>Phe (µmol/L blood)</b>	3	168.9	181.8	n/a	137.0	188.0

### International

Analyte	N	Mean	Median	Mode	Minimum	Maximum
<b>17OHP (ng/mL serum)</b>	217	25.2	20.0	19.8	2.6	100.0
<b>IRT (ng/mL blood)</b>	169	67.1	65.0	70.0	40.0	140.5
<b>T4 (µg/dL serum)</b>	49	6.5	6.0	6.0	2.4	12.8
<b>TGal (mg/dL blood)</b>	139	12.2	10.0	10.0	3.0	30.0
<b>TSH (µIU/mL serum)</b>	275	22.5	20.0	20.0	6.0	55.0
<b>Phe (µmol/L blood)</b>	27	159.7	180.0	121.2	96.8	242.4

**Table 12.** Summary of MS/MS cutoff values for domestic laboratories ( $\mu\text{mol/L}$  blood), 2018

Analyte	N	Mean	Median	Mode	Minimum	Maximum
<b>Arginine</b>	36	71.7	63.0	50.0	20.0	120.0
<b>Citrulline</b>	44	54.2	55.0	60.0	18.0	75.0
<b>Leucine</b>	44	287.9	281.5	250.0	175.0	400.0
<b>Methionine</b>	44	73.7	75.0	100.0	30.0	100.0
<b>Phenylalanine</b>	51	143.2	150.0	130.0	70.0	188.0
<b>Succinylacetone</b>	35	2.5	2.0	4.5	0.7	5.4
<b>Tyrosine</b>	48	391.6	357.5	300.0	91.0	850.0
<b>Valine</b>	30	302.0	300.0	250.0	175.0	530.0
<b>C0(L)</b>	47	8.12	7.50	6.00	4.50	24.00
<b>C3</b>	48	5.56	5.97	5.00	2.82	7.50
<b>C3DC</b>	19	0.21	0.20	0.14	0.10	0.45
<b>C3DC+ C40H</b>	21	0.52	0.38	0.38	0.25	3.03
<b>C4</b>	43	1.25	1.30	1.30	0.49	1.90
<b>C40H</b>	18	0.64	0.65	0.65	0.30	1.00
<b>C5</b>	48	0.73	0.69	0.60	0.39	1.20
<b>C5:1</b>	47	0.20	0.15	0.10	0.03	0.50
<b>C5DC</b>	47	0.37	0.39	0.50	0.05	0.80
<b>C50H</b>	47	0.80	0.80	0.80	0.25	1.36
<b>C6</b>	45	0.38	0.29	0.25	0.14	0.95
<b>C8</b>	48	0.45	0.40	0.35	0.20	0.73
<b>C10</b>	44	0.43	0.40	0.30	0.22	0.70
<b>C10:1</b>	41	0.28	0.25	0.25	0.11	0.45
<b>C10:2</b>	27	0.15	0.13	0.10	0.04	0.39
<b>C14</b>	44	0.74	0.70	0.70	0.26	1.20
<b>C14:1</b>	48	0.61	0.65	0.70	0.17	0.80
<b>C16</b>	45	7.59	7.80	10.00	2.14	10.00
<b>C160H</b>	48	0.12	0.11	0.10	0.06	0.25
<b>C18</b>	41	2.31	2.20	3.50	0.70	3.50
<b>C18:1</b>	41	3.52	3.00	2.50	2.00	7.00
<b>C180H</b>	37	0.09	0.10	0.10	0.03	0.18

**Table 13.** Summary of MS/MS cutoff values for international laboratories ( $\mu\text{mol/L}$  blood), 2018

Analyte	N	Mean	Median	Mode	Minimum	Maximum
<b>Arginine</b>	229	57.0	55.0	70.0	10.0	150.0
<b>Citrulline</b>	248	51.5	50.0	55.0	20.0	200.0
<b>Leucine</b>	272	304.3	294.5	300.0	145.0	600.0
<b>Methionine</b>	261	55.8	50.0	75.0	20.0	120.0
<b>Phenylalanine</b>	325	138.3	129.0	120.0	1.8	250.0
<b>Succinylacetone</b>	102	2.1	1.5	2.0	0.4	8.0
<b>Tyrosine</b>	269	293.6	280.0	350.0	79.9	600.0
<b>Valine</b>	254	268.3	265.0	300.0	44.0	470.0
<b>C0(L)</b>	262	16.47	9.00	10.00	2.00	100.00
<b>C3</b>	262	5.44	5.40	5.65	0.20	70.00
<b>C3DC</b>	105	0.28	0.25	0.25	0.04	1.82
<b>C3DC+ C40H</b>	100	0.49	0.44	0.45	0.15	3.07
<b>C4</b>	245	0.96	0.92	1.30	0.29	3.80
<b>C40H</b>	100	0.57	0.57	0.65	0.05	1.40
<b>C5</b>	271	0.68	0.60	0.70	0.20	2.00
<b>C5:1</b>	232	0.15	0.12	0.25	0.01	0.97
<b>C5DC</b>	256	0.33	0.30	0.35	0.07	0.90
<b>C50H</b>	226	0.76	0.79	1.00	0.19	2.50
<b>C6</b>	247	0.29	0.25	0.20	0.05	1.00
<b>C8</b>	276	0.35	0.30	0.50	0.07	1.00
<b>C10</b>	261	0.47	0.37	0.45	0.07	25.00
<b>C10:1</b>	227	0.35	0.25	0.30	0.05	20.00
<b>C10:2</b>	158	0.15	0.12	0.15	0.01	2.00
<b>C14</b>	245	0.60	0.57	0.50	0.08	1.30
<b>C14:1</b>	251	0.45	0.40	0.60	0.04	2.50
<b>C16</b>	251	6.81	7.00	7.50	0.52	14.00
<b>C160H</b>	251	0.30	0.10	0.10	0.02	48.00
<b>C18</b>	243	2.10	2.00	2.30	0.20	4.00
<b>C18:1</b>	230	3.01	3.00	3.50	0.16	8.00
<b>C180H</b>	208	0.10	0.08	0.10	0.01	2.00

**Table 14.** Domestic cutoff summary by analyte and method—hormones, galactose, and immunoreactive trypsinogen, 2018 (Methods N < 3 not shown)

### 17 $\alpha$ -Hydroxyprogesterone ng/mL serum

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>41</b>	<b>33.7</b>	<b>33.0</b>	<b>30.0</b>	<b>17.8</b>	<b>65.0</b>
AutoDelfia	6	39.6	37.5	35.0	17.8	60.0
AutoDelfia eNonatal 17-OHP (B024)	12	31.1	33.0	33.0	25.0	35.0
PerkinElmer GSP Neonatal	23	33.5	30.0	30.0	25.0	65.0

### Thyroxine $\mu$ g/dL serum

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>21</b>	<b>6.4</b>	<b>6.0</b>	<b>5.0</b>	<b>5.0</b>	<b>8.5</b>
AutoDelfia	6	6.7	6.6	n/a	5.5	8.0
PerkinElmer GSP Neonatal	14	6.3	6.0	5.0	5.0	8.5

### Thyroid-Stimulating Hormone $\mu$ IU/mL serum

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>42</b>	<b>30.7</b>	<b>25.0</b>	<b>20.0</b>	<b>19.0</b>	<b>58.0</b>
AutoDelfia	18	36.1	29.3	20.0	20.0	58.0
PerkinElmer GSP Neonatal	23	27.0	25.0	25.0	19.0	54.0

### Total Galactose mg/dL blood

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>23</b>	<b>11.2</b>	<b>10.0</b>	<b>10.0</b>	<b>6.0</b>	<b>20.0</b>
Astoria-Pacific 50 Hour Reagent Kit	4	11.5	10.5	10.0	10.0	15.0
Fluorometric manual (e.g. Hill or Misuma)	3	14.7	14.0	n/a	10.0	20.0
PerkinElmer GSP Neonatal	11	10.8	10.0	10.0	7.0	14.0

### Immunoreactive Trypsinogen ng/mL blood

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>42</b>	<b>68.3</b>	<b>63.4</b>	<b>60.0</b>	<b>46.3</b>	<b>160.0</b>
AutoDelfia	21	72.5	66.0	65.0	52.0	115.6
PerkinElmer GSP Neonatal	21	64.1	58.0	60.0	46.3	160.0

**Table 15.** Domestic cutoff summary by analyte and method—amino acids ( $\mu$ mol/L blood), 2018 (Methods N < 3 not shown)

### Arginine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>36</b>	<b>71.7</b>	<b>63.0</b>	<b>50.0</b>	<b>30.0</b>	<b>120.0</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	82.5	85.0	100.0	60.0	100.0
Derivatized - MS/MS non-kit	11	53.6	46.0	n/a.	20.0	115.0
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	19	81.4	100.0	50.0	50.0	120.0

*Continued*

## Citrulline

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>44</b>	<b>54.2</b>	<b>55.0</b>	<b>60.0</b>	<b>18.0</b>	<b>75.0</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	53.8	55.0	n/a	40.0	65.0
Derivatized - MS/MS non-kit	13	49.2	50.0	40.0	18.0	75.0
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	57.3	60.0	60.0	40.2	75.0

## Leucine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>44</b>	<b>287.9</b>	<b>281.5</b>	<b>250.0</b>	<b>175.0</b>	<b>400.0</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	281.3	275.0	275.0	250.0	325.0
Derivatized - MS/MS non-kit	13	264.6	250.0	300.0	200.0	350.0
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	303.0	300.0	250.0	175.0	400.0

## Methionine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>44</b>	<b>73.7</b>	<b>75.0</b>	<b>100.0</b>	<b>30.0</b>	<b>100.0</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	75.0	75.0	75.0	70.0	80.0
Derivatized - MS/MS non-kit	13	61.3	60.0	50.0	30.0	100.0
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	81.1	80.6	100.0	54.5	100.0

## Phenylalanine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>51</b>	<b>143.2</b>	<b>150.0</b>	<b>130.0</b>	<b>70.0</b>	<b>188.0</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	136.3	132.5	130.0	130.0	150.0
Derivatized - MS/MS non-kit	17	136.7	139.0	130.0	70.0	182.0
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	150.0	160.0	165.0	120.0	180.0
Non-derivatized - MS/MS non-kit	3	115.0	120.0	n/a	75.0	150.0

## Succinylacetone

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>35</b>	<b>2.5</b>	<b>2.0</b>	<b>4.5</b>	<b>0.7</b>	<b>5.4</b>
Derivatized - MS/MS non-kit	10	2.5	2.3	2.0	0.9	5.0
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	23	2.4	2.0	4.5	0.7	4.5

## Tyrosine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>48</b>	<b>391.6</b>	<b>357.5</b>	<b>300.0</b>	<b>91.0</b>	<b>850.0</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	282.5	290.0	300.0	250.0	300.0
Derivatized - MS/MS non-kit	16	291.9	290.0	300.0	99.0	500.0
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	23	521.3	450.0	850.0	300.0	850.0

Continued



## Valine

Method	N	Mean	Median	Mode	Min	Max
<b>ALL MS/MS METHODS</b>	<b>30</b>	<b>302.0</b>	<b>300.0</b>	<b>250.0</b>	<b>175.0</b>	<b>530.0</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	3	316.7	300.0	n/a	250.0	400.0
Derivatized - MS/MS non-kit	10	268.5	260.0	200.0	175.0	420.0
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	15	328.1	330.0	250.0	250.0	530.0

**Table 16.** Domestic cutoff summary by analyte and method—acylcarnitines (μmol/L blood), 2018 (Methods N < 3 not shown)

### C0(L)

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>47</b>	<b>8.12</b>	<b>7.50</b>	<b>6.00</b>	<b>4.50</b>	<b>24.00</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	11.12	11.23	n/a	9.00	13.00
Derivatized - MS/MS non-kit	16	9.88	8.85	10.00	5.00	24.00
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	6.48	6.00	6.00	4.50	10.00

### C3

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>48</b>	<b>5.56</b>	<b>5.97</b>	<b>5.00</b>	<b>2.82</b>	<b>7.50</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	5.54	5.57	n/a	5.00	6.00
Derivatized - MS/MS non-kit	17	4.86	5.00	5.00	2.82	7.30
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	5.96	6.30	6.30	3.09	7.50

### C3DC

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>19</b>	<b>0.21</b>	<b>0.20</b>	<b>0.14</b>	<b>0.10</b>	<b>0.45</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.25	0.26	0.30	0.19	0.30
Derivatized - MS/MS non-kit	15	0.19	0.18	0.14	0.10	0.45

### C3DC + C4OH

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>21</b>	<b>0.52</b>	<b>0.38</b>	<b>0.38</b>	<b>0.25</b>	<b>3.03</b>
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	19	0.40	0.38	0.38	0.25	0.60

### C4

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>43</b>	<b>1.25</b>	<b>1.30</b>	<b>1.30</b>	<b>0.49</b>	<b>1.90</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	1.00	0.95	n/a	0.81	1.30
Derivatized - MS/MS non-kit	15	1.14	1.20	1.40	0.49	1.90
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	22	1.39	1.32	1.70	1.06	1.70

Continued

## C4OH

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>18</b>	<b>0.64</b>	<b>0.65</b>	<b>0.65</b>	<b>0.30</b>	<b>1.00</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.73	0.68	n/a	0.55	1.00
Derivatized - MS/MS non-kit	14	0.61	0.65	0.40	0.30	1.00

## C5

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>48</b>	<b>0.73</b>	<b>0.69</b>	<b>0.60</b>	<b>0.39</b>	<b>1.20</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.66	0.70	0.70	0.54	0.70
Derivatized - MS/MS non-kit	17	0.74	0.68	0.50	0.39	1.20
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	0.74	0.70	1.00	0.50	1.00

## C5:1

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>47</b>	<b>0.20</b>	<b>0.15</b>	<b>0.10</b>	<b>0.03</b>	<b>0.50</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.20	0.20	0.15	0.15	0.25
Derivatized - MS/MS non-kit	17	0.19	0.15	0.07	0.05	0.50
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	24	0.22	0.17	0.50	0.03	0.50

## C5DC

Method	N	Mean	Median	Mode	Min	Max
<b>All Methods</b>	<b>47</b>	<b>0.37</b>	<b>0.39</b>	<b>0.50</b>	<b>0.05</b>	<b>0.80</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.28	0.28	n/a	0.24	0.32
Derivatized - MS/MS non-kit	17	0.19	0.18	0.21	0.05	0.30
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	24	0.51	0.50	0.50	0.35	0.80

## C5OH

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>47</b>	<b>0.80</b>	<b>0.80</b>	<b>0.80</b>	<b>0.25</b>	<b>1.36</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.69	0.66	n/a	0.60	0.83
Derivatized - MS/MS non-kit	17	0.79	0.80	0.80	0.25	1.36
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	24	0.81	0.83	0.85	0.60	1.05

## C6

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>45</b>	<b>0.38</b>	<b>0.29</b>	<b>0.25</b>	<b>0.14</b>	<b>0.95</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.26	0.25	0.25	0.24	0.30
Derivatized - MS/MS non-kit	16	0.33	0.31	0.24	0.14	0.63
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	23	0.45	0.26	0.95	0.16	0.95

Continued

**C8**

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>48</b>	<b>0.45</b>	<b>0.40</b>	<b>0.35</b>	<b>0.20</b>	<b>0.73</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.37	0.37	0.35	0.35	0.40
Derivatized - MS/MS non-kit	17	0.41	0.35	0.35	0.20	0.73
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	0.49	0.46	0.60	0.35	0.70

**C10**

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>44</b>	<b>0.43</b>	<b>0.40</b>	<b>0.30</b>	<b>0.22</b>	<b>0.70</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.36	0.33	n/a	0.27	0.50
Derivatized - MS/MS non-kit	15	0.38	0.40	0.30	0.22	0.55
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	23	0.48	0.45	0.65	0.22	0.70

**C10:1**

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>41</b>	<b>0.28</b>	<b>0.25</b>	<b>0.25</b>	<b>0.11</b>	<b>0.45</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.29	0.30	0.30	0.25	0.30
Derivatized - MS/MS non-kit	14	0.26	0.25	0.21	0.11	0.42
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	21	0.29	0.25	0.45	0.14	0.45

**C10:2**

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>27</b>	<b>0.15</b>	<b>0.13</b>	<b>0.10</b>	<b>0.04</b>	<b>0.39</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	3	0.17	0.15	0.15	0.15	0.20
Derivatized - MS/MS non-kit	12	0.17	0.15	0.10	0.06	0.39
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	11	0.13	0.10	0.10	0.04	0.30

**C14**

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>44</b>	<b>0.74</b>	<b>0.70</b>	<b>0.70</b>	<b>0.26</b>	<b>1.20</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.63	0.65	0.70	0.52	0.70
Derivatized - MS/MS non-kit	16	0.66	0.72	0.80	0.26	0.96
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	22	0.83	0.73	1.20	0.46	1.20

**C14:1**

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>48</b>	<b>0.61</b>	<b>0.65</b>	<b>0.70</b>	<b>0.17</b>	<b>0.80</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.55	0.55	n/a	0.40	0.70
Derivatized - MS/MS non-kit	17	0.54	0.65	0.70	0.17	0.77
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	0.68	0.68	0.80	0.50	0.80

Continued

## C16

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>45</b>	<b>7.59</b>	<b>7.80</b>	<b>10.00</b>	<b>2.14</b>	<b>10.00</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	7.04	6.93	n/a	6.50	7.80
Derivatized - MS/MS non-kit	16	6.69	7.40	8.00	2.14	9.00
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	23	8.25	8.00	10.00	5.00	10.00

## C160H

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>48</b>	<b>0.12</b>	<b>0.11</b>	<b>0.10</b>	<b>0.06</b>	<b>0.25</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.15	0.15	n/a	0.12	0.18
Derivatized - MS/MS non-kit	17	0.14	0.14	0.10	0.06	0.25
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	25	0.11	0.10	0.10	0.06	0.20

## C18

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>41</b>	<b>2.31</b>	<b>2.20</b>	<b>3.50</b>	<b>0.70</b>	<b>3.50</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	2.17	2.15	n/a	1.89	2.50
Derivatized - MS/MS non-kit	13	1.87	1.85	1.50	0.70	2.80
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	22	2.59	2.48	3.50	1.55	3.50

## C18:1

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>41</b>	<b>3.52</b>	<b>3.00</b>	<b>2.50</b>	<b>2.00</b>	<b>7.00</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	3	3.14	3.43	n/a	2.50	3.50
Derivatized - MS/MS non-kit	14	2.68	2.54	2.50	2.00	3.50
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	22	4.14	3.33	7.00	2.00	7.00

## C180H

Method	N	Mean	Median	Mode	Min	Max
<b>ALL METHODS</b>	<b>37</b>	<b>0.09</b>	<b>0.10</b>	<b>0.10</b>	<b>0.03</b>	<b>0.18</b>
Derivatized - MS/MS PerkinElmer NeoGram Kit	4	0.12	0.11	0.10	0.10	0.16
Derivatized - MS/MS non-kit	11	0.10	0.10	0.10	0.03	0.18
Non-derivatized - MS/MS PerkinElmer NeoBase Kit	20	0.09	0.10	0.10	0.03	0.16

## Explanation of the NSQAP's Grading Algorithm

NSQAP provides PT evaluations based on qualitative clinical assessments. The algorithm for determining PT errors (Figure 4) is as follows:

**Part 1:** The **NSQAP expected clinical assessment** for PT specimens is determined by comparing the **NSQAP expected value** to the **NSQAP cutoff value**.


Clinical assessments are reported as "within normal limits" or "outside normal limits." The NSQAP expected value is the sum of the endogenous value plus the enrichment value for an individual analyte. The NSQAP cutoff value is determined annually using the mean of all domestic laboratories' reported cutoff values as a guideline.

**Part 2:** The **participant reported clinical assessment** is then compared with the **NSQAP expected clinical assessment**. If these assessments agree, the algorithm stops and no error is reported. If these assessments do not agree, the grading algorithm is continued.

**Part 3:** If the algorithm was not completed in part 2, the **participant expected clinical assessment** is determined by comparing the **NSQAP expected value** to the participant's reported cutoff value. If the **participant reported clinical assessment** differs from the **participant expected clinical assessment** a false positive or false negative error will be noted. If the **participant reported clinical assessment** agrees with the **participant expected clinical assessment** a cutoff difference comment will be noted.

Determination of a final evaluation for a specimen is based on Clinical Laboratory Improvement Amendments (CLIA) regulations. These require the PT provider to compare the laboratory's response for each analyte with the response that reflects agreement of 80% or more of all laboratories. (CLIA Regulations, 2004). An NSQAP gradable specimen must have 80% or more agreement among domestic laboratories. For analytes with less than 10 domestic participants, the specimen will be evaluated unless the sample is deemed ungradable by the review committee.

Figure 2. NSQAP's Grading Algorithm Flow chart





# 2018 Bias Plots

## Proficiency Testing Bias Plots

Figures 3–36 are illustrated for PT analytes reported using the NSQAP data reporting website. A wide range of quantitatively measured PT challenges was selected for the bias plots. Comparisons of results by different methods are illustrated with the participants' reported PT data for one selected challenge for each analyte. The expected value of each specimen equals the sum of the enriched value and the endogenous (non-enriched) value. Immunoreactive trypsinogen (IRT) standard cannot be fully recovered by any IRT analytical method; therefore, IRT PT uses CDC-assayed values.

Non-derivatized MS/MS methods for amino acids and acylcarnitine analysis cannot distinguish between analytes C3DC and C4OH (i.e., they are isobaric). Laboratories using a non-derivatized MS/MS method report C3DC+C4OH, while derivatized MS/MS method users report those analytes separately. These bias plots show the difference of the reported value (positive or negative) by laboratory and method subtracted from the

expected or assayed value. To illustrate method-related differences in analyte recoveries, the PT quantitative results are grouped by kit or method.


For each plot, note the scale-changes of the y-axis. A reported value matching the expected value (endogenous value plus enriched value) falls on the plot's "0" line. For each figure, a summary of the specimen data for the selected PT challenge is tabulated in the left margin. Ideally, a reasonable bias is less than 20% of the expected value.

The bias plots illustrate the 95% confidence interval for the participant mean. A tight scatter within this interval indicates good performance for a method or a group of methods. In general, the quantitative comparisons for PT challenges are reasonable within a method but vary among methods. Because some of the pools in a routine PT survey represent a unique donor specimen, differences in endogenous materials in the donor specimens might influence method-related differences.

Note for accessibility:

For Figures 3–36, the bias plot's explanation follows each figure title.

**Figure 3. Reproducibility of Results:  
Bias Plot of 17  $\alpha$ -Hydroxyprogesterone (17OHP) Values by Method  
Quarter 1, Specimen 11815  
Expected Value (EV) = 86.0 ng/mL serum**



The 17OHP bias plot shows units of measure on the y-axis ranging from 150 ng/mL serum to -150 ng/mL serum. The mean bias for this plot is -6.7 ng/mL serum. The data on this plot shows a tight scatter among all participants.

**17OHP** ng/mL serum

**Quarter 1**


Enriched—85.0

CDC Assayed—71.4

Participant Mean—79.3

Participant Bias—-6.7

**Figure 4. Reproducibility of Results:  
Thyroxine (T4) Values by Method  
Quarter 1, Specimen 11811  
Expected Value (EV) = 1.6  $\mu$ g/dL serum**



The T4 bias plot shows units of measure on the y-axis ranging from 2  $\mu$ g/dL serum to -2  $\mu$ g/dL serum. The mean bias for this plot is zero. The data on this plot shows a good agreement among participants.

**T4**  $\mu$ g/dL serum

**Quarter 1**


Enriched—1.5

CDC Assayed—1.5

Participant Mean—1.6


Participant Bias—0.0

**Figure 5. Reproducibility of Results:  
Bias Plot for Thyroid-Stimulating Hormone (TSH) Values by Method  
Quarter 1, Specimen 11811  
Expected Value (EV) = 85.6  $\mu$ U/mL serum**



The TSH bias plot shows units of measure on the y-axis ranging from 100  $\mu$ U/mL serum to -100  $\mu$ U/mL serum. The mean bias for this plot is 12.1  $\mu$ g/dL. This plot shows a positive bias compared to the CDC-expected value. All methods show a tight scatter with most participants clustering in a positive bias.

**Figure 6. Reproducibility of Results:  
Bias Plot for Total Galactose (TGal) Values by Method  
Quarter 1, Specimen 11814  
Expected Value (EV) = 25.4 mg/dL blood**



The TGal bias plot shows units of measure on the y-axis ranging from 50 mg/dL blood to -50 mg/dL blood. This plot shows a slight positive bias compared to the CDC-expected value. There is good agreement within each method however, some methods show a positive bias while other methods show a negative bias.

**TSH**  $\mu$ U/mL serum

**Quarter 1**


- Enriched—85.0
- CDC Assayed—104.9
- Participant Mean—97.7
- Participant Bias—12.1

**TGal** mg/dL blood

**Quarter 1**

- Enriched—25.0
- CDC Assayed—20.3
- Participant Mean—25.7
- Participant Bias—0.3

**Figure 7. Reproducibility of Results:  
Bias Plot for Immunoreactive Trypsinogen (IRT) Values by Method  
Quarter 1, Specimen 11882  
Expected Value (EV) = 256.5 ng/mL blood**




The IRT bias plot shows units of measure on the y-axis ranging from 300 ng/mL blood to -300 ng/mL blood. The mean bias for this plot is -32.4 ng/mL blood. There is tight scatter with all except for the other method category which shows a wide range of results.

**IRT** ng/mL blood

**Quarter 1**

- Enriched**—400.0
- CDC Assayed**—256.5
- Participant Mean**—224.1
- Participant Bias**—-32.4

**Figure 8. Reproducibility of Results:  
Bias Plot of Arginine (Arg) Values by Method  
Quarter 1, Specimen 11853  
Expected Value (EV) = 258.6 μmol/L blood**




The Arg bias plot shows units of measure on the y-axis ranging from 200 μmol/L blood to -300 μmol/L blood. The mean bias for this plot is -110.1 μmol/L blood. When compared to the CDC expected value, this plot shows a negative bias for all methods.

**Arg** μmol/L blood

**Quarter 1**


- Enriched**—250.0
- CDC Assayed**—198.7
- Participant Mean**—148.5
- Participant Bias**—-110.1

**Figure 9. Reproducibility of Results:  
Bias Plot for Citrulline (Cit) Values by Method  
Quarter 3, Specimen 31852  
Expected Value (EV) = 251.2  $\mu\text{mol/L}$  blood**



The Cit bias plot shows units of measure on the y-axis ranging from 300  $\mu\text{mol/L}$  blood to -300  $\mu\text{mol/L}$  blood. The mean bias for this plot is -10.39  $\mu\text{mol/L}$  blood. The Cit bias plot shows a good agreement among methods.

**Figure 10. Reproducibility of Results:  
Bias Plot for Leucine (Leu) Values by Method  
Quarter 3, Specimen 31854  
Expected Value (EV) = 773.8  $\mu\text{mol/L}$  blood**



The Leu bias plot shows units of measure on the y-axis ranging from 700  $\mu\text{mol/L}$  blood to -700  $\mu\text{mol/L}$  blood. The mean bias for this plot is -172.5  $\mu\text{mol/L}$  blood. The bias plot shows distinct differences between methods with some methods below the bias and some above.

**Cit**  $\mu\text{mol/L}$  blood

**Quarter 3**

Enriched—239.2

CDC Assayed—230.3

Participant Mean—240.8

Participant Bias—-10.39

**Leu**  $\mu\text{mol/L}$  blood

**Quarter 3**

Enriched—748.8

CDC Assayed—770.3

Participant Mean—601.3

Participant Bias—-172.5



**Figure 11. Reproducibility of Results:  
Bias Plot for Methionine (Met) Values by Method  
Quarter 3, Specimen 31851  
Expected Value (EV) = 204.4  $\mu\text{mol/L}$  blood**

**Met**  $\mu\text{mol/L}$  blood


**Quarter 3**

Enriched—200.4

CDC Assayed—195.6

Participant Mean—194.8

Participant Bias—9.6



The Met bias plot shows units of measure on the y-axis ranging from 200  $\mu\text{mol/L}$  blood to -200  $\mu\text{mol/L}$  blood. The mean bias for this plot is -9.6  $\mu\text{mol/L}$  blood. The data shows good scatter among all methods.

**Figure 12. Reproducibility of Results:  
Phenylalanine (Phe) Values by Method  
Quarter 1, Specimen 11851  
Expected Value (EV) = 285.0  $\mu\text{mol/L}$  blood**

**Phe**  $\mu\text{mol/L}$  blood


**Quarter 1**

Enriched—225.0

CDC Assayed—250.9

Participant Mean—257.0

Participant Bias—28.0




The Phe bias plot shows units of measure on the y-axis ranging from 300  $\mu\text{mol/L}$  blood to -300  $\mu\text{mol/L}$  blood. The bias for this plot is -28.0  $\mu\text{mol/L}$  blood. The Phe plot shows good agreement between and among methods.

**Figure 13. Reproducibility of Results:  
Succinylacetone (SUAC) Values by Method  
Quarter 1, Specimen 11854  
Expected Value (EV) = 15.4  $\mu\text{mol/L}$  blood**

**SUAC**  $\mu\text{mol/L}$  blood

**Quarter 1**

- Enriched—15.0
- CDC Assayed—6.3
- Participant Mean—5.7
- Participant Bias—9.7




The SUAC bias plot shows units of measure on the y-axis ranging from 20  $\mu\text{mol/L}$  blood to -20  $\mu\text{mol/L}$  blood. The mean bias for this plot is -9.7  $\mu\text{mol/L}$  blood. The SUAC bias plot shows a very tight scatter among methods with only a few outliers.

**Figure 14. Reproducibility of Results:  
Bias Plot for Tyrosine (Tyr) Values by Method  
Quarter 1, Specimen 11854  
Expected Value (EV) = 731.8  $\mu\text{mol/L}$  blood**

**Tyr**  $\mu\text{mol/L}$  blood

**Quarter 1**

- Enriched—675.0
- CDC Assayed—617.6
- Participant Mean—626.5
- Participant Bias—105.3




The Tyr bias plot shows units of measure on the y-axis ranging from 400  $\mu\text{mol/L}$  blood to -400  $\mu\text{mol/L}$  blood. The mean bias for this plot is -105.3  $\mu\text{mol/L}$  blood. The bias plot shows good scatter among participants and methods.

**Figure 15. Reproducibility of Results:  
Bias Plot for Valine (Val) Values by Method  
Quarter 3, Specimen 31854  
Expected Value (EV) = 629.1  $\mu\text{mol/L}$  blood**

**Val**  $\mu\text{mol/L}$  blood

**Quarter 3**

Enriched—604.1  
 CDC Assayed—646.3  
 Participant Mean—624.5  
 Participant Bias—-4.61




The Val bias plot shows units of measure on the y-axis ranging from 400  $\mu\text{mol/L}$  blood to -400  $\mu\text{mol/L}$  blood. The mean bias for this plot is -4.61  $\mu\text{mol/L}$  blood. The Val bias plot shows good scatter among all participants and methods.

**Figure 16. Reproducibility of Results:  
Bias Plot of Free Carnitine(C0(L)) Values by Method  
Quarter 3, Specimen 31865  
Expected Value (EV) = 41.51  $\mu\text{mol/L}$  blood**

**C0(L)**  $\mu\text{mol/L}$  blood

**Quarter 3**

Enriched—31.56  
 CDC Assayed—46.16  
 Participant Mean—40.97  
 Participant Bias—-0.54




The C0(L) bias plot shows units of measure on the y-axis ranging from 55  $\mu\text{mol/L}$  blood to -55  $\mu\text{mol/L}$  blood. The mean bias for this plot is -0.54  $\mu\text{mol/L}$  blood. The bias for this specimen is very close to the CDC expected value and there is tight scatter among all methods.

**Figure 17. Reproducibility of Results:  
Bias Plot of Propionylcarnitine (C3) Values by Method  
Quarter 1, Specimen 11865  
Expected Value (EV) = 11.69  $\mu\text{mol/L}$  blood**

**C3**  $\mu\text{mol/L}$  blood

**Quarter 1**

Enriched—11.00  
 CDC Assayed—12.80  
 Participant Mean—10.22  
 Participant Bias—-1.47




The C3 bias plot shows units of measure on the y-axis ranging from 15  $\mu\text{mol/L}$  blood to -15  $\mu\text{mol/L}$  blood. The mean bias for this plot is -1.47  $\mu\text{mol/L}$  blood. The bias plot shows good scatter around the bias.

**Figure 18. Reproducibility of Results:  
Bias Plot of Malonylcarnitine (C3DC) Values by Method  
Quarter 4, Specimen 41862  
Expected Value (EV) = 0.03  $\mu\text{mol/L}$  blood**

**C3DC**  $\mu\text{mol/L}$  blood

**Quarter 4**

Enriched—0.00  
 CDC Assayed—0.03  
 Participant Mean—0.04  
 Participant Bias—0.01



The C3DC bias plot shows units of measure on the y-axis ranging from 0.5  $\mu\text{mol/L}$  blood to -0.5  $\mu\text{mol/L}$  blood. The mean bias for this plot is 0.01  $\mu\text{mol/L}$  blood. The expected value of this specimen was very close to zero. The participant bias was very close to the expected value and with the exception of a few outliers, all data points are very close to the bias.

**Figure 19. Reproducibility of Results:  
Bias Plot of Butylcarnitine (C4) Values by Method  
Quarter 1, Specimen 11861  
Expected Value (EV) = 0.08  $\mu\text{mol/L}$  blood**

**C4**  $\mu\text{mol/L}$  blood


**Quarter 1**

Enriched—0.00

CDC Assayed—0.09

Participant Mean—0.10

Participant Bias—0.02



The C4 bias plot shows units of measure on the y-axis ranging from 0.5  $\mu\text{mol/L}$  blood to -0.5  $\mu\text{mol/L}$  blood. The mean bias for this plot is 0.02  $\mu\text{mol/L}$  blood. All methods show a very tight scatter very close to the bias.

**Figure 20. Reproducibility of Results:  
Bias Plot for Hydroxybutyrylcarnitine (C4OH) Values by Method  
Quarter 4, Specimen 41862  
Expected Value (EV) = 0.20  $\mu\text{mol/L}$  blood**

**C4OH**  $\mu\text{mol/L}$  blood


**Quarter 4**

Enriched—0.00

CDC Assayed—0.20

Participant Mean—0.18

Participant Bias—-0.2




The C4 bias plot shows units of measure on the y-axis ranging from 0.5  $\mu\text{mol/L}$  blood to -0.5  $\mu\text{mol/L}$  blood. The bias for this plot is -0.02  $\mu\text{mol/L}$  blood. All methods show a very tight scatter very close to the bias.

**Figure 21. Reproducibility of Results:  
Bias Plot for Isovalerylcarnitine (C5) Values by Method  
Quarter 3, Specimen 31863  
Expected Value (EV) = 2.99  $\mu\text{mol/L}$  blood**

**C5**  $\mu\text{mol/L}$  blood

**Quarter 3**

- Enriched—2.92
- CDC Assayed—3.13
- Participant Mean—2.78
- Participant Bias—-0.21




The C5 bias plot shows units of measure on the y-axis ranging from 2  $\mu\text{mol/L}$  blood to -2  $\mu\text{mol/L}$  blood. The mean bias for this plot is -0.21  $\mu\text{mol/L}$  blood. The C5 plot shows a good scatter close to the bias for all methods.

**Figure 22. Reproducibility of Results:  
Bias Plot for Tiglylcarnitine (C5:1) Values by Method  
Quarter 1, Specimen 11861  
Expected Value (EV) = 0.01  $\mu\text{mol/L}$  blood**

**C5:1**  $\mu\text{mol/L}$  blood

**Quarter 1**


- Enriched—0.00
- CDC Assayed—0.01
- Participant Mean—0.02
- Participant Bias—0.01



The C5:1 bias plot shows units of measure on the y-axis ranging from 0.3  $\mu\text{mol/L}$  blood to -0.3  $\mu\text{mol/L}$  blood. The mean bias for this plot is 0.01  $\mu\text{mol/L}$  blood. The expected value of this specimen was very close to zero. The participant bias was very close the to expected value and with the exception of a few outliers, all data points are very close to the bias.




**Figure 23. Reproducibility of Results:  
Bias Plot for Glutarylcarnitine (C5DC) Values by Method  
Quarter 1, Specimen 11862  
Expected Value (EV) = 1.31  $\mu\text{mol/L}$  blood**



The C5DC bias plot shows units of measure on the y-axis ranging from 2  $\mu\text{mol/L}$  blood to -2  $\mu\text{mol/L}$  blood. The mean bias for this plot is -0.06  $\mu\text{mol/L}$  blood. The C5DC plot shows a tight scatter within each method but shows distinctive bias differences between methods.

**Figure 24. Reproducibility of Results:  
Bias Plot for Hydroxyisovalerylcarnitine (C5OH) Values by Method  
Quarter 3, Specimen 31864  
Expected Value (EV) = 1.83  $\mu\text{mol/L}$  blood**



The C5OH bias plot shows units of measure on the y-axis ranging from 2  $\mu\text{mol/L}$  blood to -2  $\mu\text{mol/L}$  blood. The bias for this plot is -0.51  $\mu\text{mol/L}$  blood. The C5DC plot shows a slight negative bias but good scatter among most methods.

**C5DC**  $\mu\text{mol/L}$  blood

**Quarter 1**

Enriched—1.30

CDC Assayed—1.30

Participant Mean—1.25

Participant Bias—-0.06

**C5OH**  $\mu\text{mol/L}$  blood

**Quarter 3**

Enriched—1.29

CDC Assayed—1.53

Participant Mean—1.32

Participant Bias—-0.51

**Figure 25. Reproducibility of Results:  
Bias Plot for Hexanoylcarnitine (C6) Values by Method  
Quarter 3, Specimen 31862  
Expected Value (EV) = 1.54  $\mu\text{mol/L}$  blood**

**C6**  $\mu\text{mol/L}$  blood


**Quarter 3**

Enriched—1.53

CDC Assayed—1.33

Participant Mean—1.31

Participant Bias—-0.23



The C6 bias plot shows units of measure on the y-axis ranging from 2  $\mu\text{mol/L}$  blood to -2  $\mu\text{mol/L}$  blood. The bias for this plot is -0.23  $\mu\text{mol/L}$  blood. The plot shows a negative participant bias with good scatter among methods.

**Figure 26. Reproducibility of Results:  
Bias Plot for Octanoylcarnitine (C8) Values by Method  
Quarter 1, Specimen 11861  
Expected Value (EV) = 1.63  $\mu\text{mol/L}$  blood**

**C8**  $\mu\text{mol/L}$  blood


**Quarter 1**

Enriched—1.60

CDC Assayed—1.66


Participant Mean—1.57

Participant Bias—-0.06



The C8 bias plot shows units of measure on the y-axis ranging from 2  $\mu\text{mol/L}$  blood to -2  $\mu\text{mol/L}$  blood. The bias for this plot is -0.06  $\mu\text{mol/L}$  blood. The participant bias is very close to the CDC expected value and there is good scatter among methods.

**Figure 27. Reproducibility of Results:  
Bias Plot for Decanoylcarnitine (C10) Values by Method  
Quarter 1, Specimen 11861  
Expected Value (EV) = 0.81  $\mu\text{mol/L}$  blood**



The C10 bias plot shows units of measure on the y-axis ranging from 2  $\mu\text{mol/L}$  blood to -2  $\mu\text{mol/L}$  blood. The bias for this plot is -0.07  $\mu\text{mol/L}$  blood. One method shows a distinct negative bias. The other methods show good scatter around the bias.

**C10**  $\mu\text{mol/L}$  blood

**Quarter 1**


Enriched—0.00

CDC Assayed—0.80

Participant Mean—0.74

Participant Bias—-0.07

**Figure 28. Reproducibility of Results:  
Bias Plot for Decenoylcarnitine (C10:1) Values by Method  
Quarter 1, Specimen 11861  
Expected Value (EV) = 0.52  $\mu\text{mol/L}$  blood**



The C10:1 bias plot shows units of measure on the y-axis ranging from 1  $\mu\text{mol/L}$  blood to -1  $\mu\text{mol/L}$  blood. The bias for this plot is -0.03  $\mu\text{mol/L}$  blood. On the C10:1 bias plot, there is good agreement within methods but some methods show a positive bias and others show a negative bias.

**C10:1**  $\mu\text{mol/L}$  blood

**Quarter 1**


Enriched—0.50

CDC Assayed—0.58

Participant Mean—0.49


Participant Bias—-0.03

**Figure 29. Reproducibility of Results:  
Bias Plot of Decadienoylcarnitine (C10:2) Values by Method  
Quarter 3, Specimen 31862  
Expected Value (EV) = 0.00  $\mu\text{mol/L}$  blood**



The C10:2 bias plot shows units of measure on the y-axis ranging from 0.5  $\mu\text{mol/L}$  blood to -0.5  $\mu\text{mol/L}$  blood. The bias for this plot is 0.01  $\mu\text{mol/L}$  blood. The expected value of this specimen was very close to zero and the participant bias was very close to the expected value. With the exception of a few outliers, all data points are very close to the bias.

**Figure 30. Reproducibility of Results:  
Bias Plot for Myristoylcarnitine (C14) Values by Method  
Quarter 1, Specimen 11864  
Expected Value (EV) = 1.34  $\mu\text{mol/L}$  blood**



The C14 bias plot shows units of measure on the y-axis ranging from 1  $\mu\text{mol/L}$  blood to -1  $\mu\text{mol/L}$  blood. The bias for this plot is -0.23  $\mu\text{mol/L}$  blood. The C14 plot shows a slight negative bias and the majority of data points are below the bias.

**C10:2**  $\mu\text{mol/L}$  blood

**Quarter 3**

Enriched—0.00

CDC Assayed—0.00

Participant Mean—0.01

Participant Bias—0.01

**C14**  $\mu\text{mol/L}$  blood

**Quarter 1**

Enriched—1.30

CDC Assayed—1.18

Participant Mean—1.11


Participant Bias—0.23

**Figure 31. Reproducibility of Results:  
Bias Plot of Tetradecenoylcarnitine (C14:1) Values by Method  
Quarter 4, Specimen 41861  
Expected Value (EV) = 1.31  $\mu\text{mol/L}$  blood**

**C14:1  $\mu\text{mol/L}$  blood**

**Quarter 4**

- Enriched—1.28
- CDC Assayed—0.78
- Participant Mean—0.64
- Participant Bias—-0.67




The C14:1 bias plot shows units of measure on the y-axis ranging from 1.5  $\mu\text{mol/L}$  blood to -1.5  $\mu\text{mol/L}$  blood. The bias for this plot is -0.67  $\mu\text{mol/L}$  blood. The non-kit methods show a tight negative cluster around the bias and the kit methods are scatter above and below the bias.

**Figure 32. Reproducibility of Results:  
Palmitoylcarnitine (C16) Values by Method  
Quarter 3, Specimen 31865  
Expected Value (EV) = 3.70  $\mu\text{mol/L}$  blood**

**C16  $\mu\text{mol/L}$  blood**

**Quarter 3**

- Enriched—3.24
- CDC Assayed—3.75
- Participant Mean—3.75
- Participant Bias—0.05




The C16 bias plot shows units of measure on the y-axis ranging from 4  $\mu\text{mol/L}$  blood to -4  $\mu\text{mol/L}$  blood. The bias for this plot is 0.05  $\mu\text{mol/L}$  blood. The C16 bias shows a bias very close to zero with good scatter among all participants and methods.

**Figure 33. Reproducibility of Results:  
Hydroxypalmitoycarnitine (C16OH) Values by Method  
Quarter 4, Specimen 41861  
Expected Value (EV) = 1.40  $\mu\text{mol/L}$  blood**

**C16OH**  $\mu\text{mol/L}$  blood

**Quarter 4**

- Enriched—1.39
- CDC Assayed—1.39
- Participant Mean—0.92
- Participant Bias—-0.48




The C16OH bias plot shows units of measure on the y-axis ranging from 1.5  $\mu\text{mol/L}$  blood to -1.5  $\mu\text{mol/L}$  blood. The bias for this plot is -0.48  $\mu\text{mol/L}$  blood. The C16OH bias plot demonstrates scatter among all methods with most laboratories showing a negative bias.

**Figure 34. Reproducibility of Results:  
Bias Plot for Stearoylcarnitine (C18) Values by Method  
Quarter 3, Specimen 31865  
Expected Value (EV) = 1.73  $\mu\text{mol/L}$  blood**

**C18**  $\mu\text{mol/L}$  blood

**Quarter 3**


- Enriched—1.29
- CDC Assayed—1.37
- Participant Mean—1.40
- Participant Bias—-0.33



The C18 bias plot shows units of measure on the y-axis ranging from 1.5  $\mu\text{mol/L}$  blood to -1.5  $\mu\text{mol/L}$  blood. The bias for this plot is -0.33  $\mu\text{mol/L}$  blood. The C18 bias plot shows reasonable scatter of values within and among methods while showing a slight negative bias.




**Figure 35. Reproducibility of Results:  
Bias Plots for Oleoylcarnitine (C18:1) Values by Method  
Quarter 1, Specimen 11865  
Expected Value (EV) = 0.84  $\mu\text{mol/L}$  blood**



The C18 bias plot shows units of measure on the y-axis ranging from 1  $\mu\text{mol/L}$  blood to -1  $\mu\text{mol/L}$  blood. The mean bias for this plot is -0.06  $\mu\text{mol/L}$  blood. The C18 bias plot shows reasonable scatter of values within and among methods while showing a slight negative bias.

**Figure 36. Reproducibility of Results:  
Bias Plot of Hydroxystearoylcarnitine (C18OH) Values by Method  
Quarter 4, Specimen 41861  
Expected Value (EV) = 1.47  $\mu\text{mol/L}$  blood**



The C18OH bias plot shows units of measure on the y-axis ranging from 2  $\mu\text{mol/L}$  blood to -2  $\mu\text{mol/L}$  blood. The bias for this plot is -0.64  $\mu\text{mol/L}$  blood. The C18OH plot shows a negative bias with all methods clustered around the bias.

**C18:1**  $\mu\text{mol/L}$  blood

**Quarter 1**

- Enriched—0.00
- CDC Assayed—0.85
- Participant Mean—0.78
- Participant Bias—-0.06

**C18OH**  $\mu\text{mol/L}$  blood

**Quarter 4**

- Enriched—1.46
- CDC Assayed—0.85
- Participant Mean—0.83
- Participant Bias—-0.64

# Appendix for Accessibility Descriptions

**Figure 2:** NSQAP’s Grading Algorithm Flow chart.

1. PART 1 is in a square box and makes the statement, “COMPARE NSQAP EXPECTED VALUE TO NSQAP CUTOFF VALUE TO DETERMINE NSQAP EXPECTED CLINICAL ASSESSMENT”.
2. A down arrow points to an oval shape and asks the question, “DOES PARTICIPANT REPORTED CLINICAL ASSESSMENT DIFFER FROM NSQAP EXPECTED CLINICAL ASSESSMENT?”
3. A right side arrow from the oval points to a smaller oval with the statement, “IF “NO:” NO ERROR”
4. A down arrow from the oval contains a solid oval within it, and the words, “IF ‘YES’”. The down arrow points to PART 2 in a square box that says “PART 2 COMPARE NSQAP EXPECTED VALUE TO PARTICIPANT REPORTED CUTOFF VALUE TO DETERMINE PARTICIPANT EXPECTED CLINICAL ASSESSMENT”
5. A down arrow points to PART 3 in an oval shape and asks the question, “DOES PARTICIPANT REPORTED CLINICAL ASSESSMENT DIFFER FROM PARTICIPANT EXPECTED CLINICAL ASSESSMENT?”
6. A right side arrow from the oval points to a smaller oval with the statement, “IF “NO:” CUTOFF DIFFERENCE COMMENT”

**Figures 5–38, Bias Plots:** Bias plots, which compare two measurements of the same variable, have been created to show a wide range of PT challenge specimens. The bias, which is calculated by subtracting the participant mean value from the CDC Expected Value (EV), is represented by the broken line. Expected Value is the sum of the endogenous plus the enrichment values. The solid line represents perfect agreement with the EV or zero bias. When comparing data scatter among figures, the scale (y-axis) might differ. We included the 95% confidence interval for the mean participant bias. A tight scatter within this interval indicates good performance for a method or a group of methods. To illustrate any method-related differences in analyte recoveries, we group the PT quantitative results by kit or method. Because some of the pools in a routine PT survey represent a unique donor specimen, differences in endogenous materials in the donor specimens might influence method-related differences. We show representative bias plots for all those analytes distributed in PT challenges that required a quantitative measurement to determine the presumptive clinical assessments.

## References

- [1] Newborn Screening: Towards a Uniform Screening Panel and System.” Genetic Medicine 2006;8(5) Suppl: S12–S252, as authored by the American College of Medical Genetics and commissioned by the Health Resources and Services Administration.
- [2] De Jesús VR, Mei JV, Cordovado SK, Cuthbert CD. The Newborn Screening Quality Assurance Program at the Centers for Disease Control and Prevention: Thirty-Five Year Experience Assuring Newborn Screening Laboratory Quality. International Journal of Newborn Screening 2015;1; 13-26.
- [3] Clinical and Laboratory Standards Institute. Blood collection on filter paper for newborn screening programs: Approved Standard—Sixth Edition. CLSI Document NBS01-A6. Wayne, PA: Clinical and Laboratory Standards Institute; 2014.

# Acknowledgments

This NEWBORN SCREENING QUALITY ASSURANCE PROGRAM report is an internal publication distributed to program participants and selected program colleagues. The laboratory quality assurance program is a project cosponsored by the Centers for Disease Control and Prevention (CDC) and the Association of Public Health Laboratories.

## CENTERS FOR DISEASE CONTROL AND PREVENTION (CDC) ATLANTA, GA 30341

### Director

Robert R. Redfield, M.D.

### Director

National Center for Environmental Health  
Patrick Breyse, Ph.D.

### Director

Division of Laboratory Sciences  
James L. Pirkle, M.D., Ph.D.

### Chief

Newborn Screening and Molecular Biology Branch  
Carla Cuthbert, Ph.D.

---

### Contributors

Carter Asef, BS

Nicole Baird, Ph.D

John Bernstein, MS

Quan Bui, MS

Suzanne Cordovado, Ph.D

Paul Dantonio, MS

Katherine Duneman, MS

Sharon Flores, MS

Christopher Greene, Ph.D

Elizabeth Hall, BS

Laura Hancock, MS

Christopher Haynes, Ph.D

Jessica Hendricks, MS

Miyono Hendrix, MS

Laura C. Hildreth, BS

Deborah Koontz, Ph.D

Francis Lee, Ph.D

LiXia Li, Ph.D

Tim Lim, Ph.D

Daniel Mandel, Ph.D

Joanne Mei, Ph.D

Kristina Mercer, Ph.D

Stanimila Nikolova, Ph.D

Gyliann Pena, MS

Kostas Petritis, Ph.D

C. Austin Pickens, Ph.D

Blanche Temate, Ph.D

E. Shannon Torres, Ph.D

Robert Vogt, Ph.D

Irene Williams, MS

Sophia Winchester, BS

Golriz Yazdanpanah, MS

Sherri Zobel, BS

### Production

Vinay Anumula, MS

Kizzy Stewart

Joy Pressley

---

## ASSOCIATION OF PUBLIC HEALTH LABORATORIES SILVER SPRING, MD 20910

### President

Joanne Bartkus, PhD

### Chairman, Newborn Screening and Genetics in Public Health Committee

Michele Caggana, Sc.D., FACMG

### Chairman, Newborn Screening Quality Assurance Quality Control Subcommittee

Patricia R. Hunt, B.A. and Joseph Orsini, Ph.D.

### Chairman, Newborn Screening Molecular Subcommittee

Rachel Lee, Ph.D.

---

### INQUIRIES TO:

Sherri Zobel, Editor

Centers for Disease Control and Prevention (CDC),

Newborn Screening Quality Assurance Program—Mailstop F-19

4770 Buford Highway, N.E., Atlanta, GA 30341-3724

E-mail: [NSQAPDMT@cdc.gov](mailto:NSQAPDMT@cdc.gov)



**For more information please contact**

Centers for Disease Control and Prevention  
1600 Clifton Road NE, Atlanta, GA 33029-4027

Telephone: 1-800-CDC-INFO (232-4636)

TTY: 1-888-232-6348

E-mail: [NSQAPDMT@cdc.gov](mailto:NSQAPDMT@cdc.gov)

Web: <https://www.cdc.gov/labstandards/nsgap.html>

Publication date: September 2019