

## X-linked Adrenoleukodystrophy in Dried Blood Spots Proficiency Testing Program (XALDPT)

Date issued: February 28, 2018

### Introduction

This is a summary of data reported within the specified data-reporting period for Quarter 1, 2018, for the detection of X-ALD by analysis of the biomarkers 24:0-Lysophosphatidylcholine (24LPC) and 26:0-Lysophosphatidylcholine (26LPC) in dried blood spots (DBS). It is distributed to all participants, state laboratory directors, and program colleagues by request. The tables within this report provide certification profiles for the distributed specimens, statistical analysis of participant quantitative data, and frequency of clinical assessments. An evaluation of your laboratory's data is attached to this summary.

### Certification of PT Specimens

This panel of DBS specimens was prepared from Type A+ human whole blood, which was adjusted to a hematocrit of  $50 \pm 1\%$  and subsequently enriched with the biomarkers 24LPC and 26LPC. Expected values for each were determined by LC-MS/MS in units of  $\mu\text{mol/L}$  blood. Clinical assessments were based on the NSQAP cut-offs of  $0.47 \mu\text{mol/L}$  blood (24LPC) and  $0.39 \mu\text{mol/L}$  blood (26LPC). Table 1 shows the NSQAP expected values and clinical assessments for each specimen.

Table 1. Specimen Certification – 24LPC and 26LPC ( $\mu\text{mol/L}$  blood)

Specimen	Expected 24LPC	24LPC Assessment Code*	Expected 26LPC	26LPC Assessment Code*
11821	1.07	2	1.04	2
11822	0.07	1	0.04	1
11823	0.07	1	0.04	1
11824	0.07	1	0.04	1
11825	0.07	1	0.04	1

\*1 = Within Normal Limits  
2 = Outside Normal Limits

## Distribution of PT Specimens

On January 9, 2018 a PT panel of five unknown DBS specimens was distributed to 11 domestic laboratories and 13 foreign laboratories.

## Participant Results

### ◆ Quantitative Data

We processed data from 16 participants, with one participant submitting two method assessments. Laboratories were asked to report concentrations of 24LPC and 26LPC results in  $\mu\text{mol/L}$  blood. In order to expedite the issuance of this report, data that are not submitted in the requested units are not accepted. The conversion factor from  $\mu\text{g/mL}$  to  $\mu\text{mol/L}$  blood is provided on the XALDPT Data Report Form. Participants may contact us for guidance on conversion factors if needed.

Overall statistics from MS/MS methods were combined so as to not identify an individual laboratory. We also did not include data that were outside the 99% confidence interval. The statistical summary analysis for all methods is provided in Table 2.

Six participants reported using Flow Injection Analysis (FIA) MS/MS non-kit, ten reported using LC-MS/MS and one reported a two-tier assessment scheme utilizing both FIA- and LC-MS/MS. There were four submissions of quantitative results for 24LPC, two without reporting a clinical assessment. There were also 17 submissions of quantitative results and clinical assessments for 26LPC. One participant reported cutoffs for 24LPC using multivariate analysis by a post-analytic tool and 2nd tier testing when indicated. Table 2b shows the reported cutoffs for 24LPC and 26LPC by reported method.

Table 2. Screening Results for 24LPC and 26LPC — All MS/MS methods

Analyte	Specimen	N	Mean ( $\mu\text{mol/L}$ )	SD
24LPC	11821	16	1.17	0.37
	11822	16	0.13	0.09
	11823	16	0.12	0.08
	11824	16	0.13	0.08
	11825	16	0.13	0.08
26LPC	11821	17	1.23	0.25
	11822	17	0.13	0.13
	11823	17	0.12	0.13
	11824	17	0.12	0.12
	11825	17	0.11	0.12

Table 2b. Reported Cutoffs by Reported Method ( $\mu\text{mol/L}$ )

	53– LC-MS/MS		67 FIA-MS/MS	
	24LPC	26LPC	24LPC	26LPC
N	8	11	2	4
Mean	0.39	0.34	0.48	0.49
Max	0.55	0.75	0.62	0.65
Min	0.16	0.10	0.33	0.40
Median	0.44	0.30	0.48	0.45
Mode	0.50	0.40	NA	NA

◆ Clinical Assessments

Laboratories were asked to report qualitative results as “Within Normal Limits” or “Outside Normal Limits”. Qualitative assessments may differ because of specific assessment practices. The frequency distribution of participants’ clinical assessments is shown in Table 3.

Table 3. Frequency Distribution of reported Clinical Assessments

Analyte	Specimen	Within Normal Limits	Outside Normal Limits
24LPC	11821	0	12
	11822	12	0
	11823	12	0
	11824	12	0
	11825	12	0
26LPC*	11821	0	17
	11822	17	0
	11823	17	0
	11824	17	0
	11825	17	0

\*One participant submitted clinical assessments for two methods (First- and Second-tier).

**Evaluations**

No False-negatives and no False-positives were reported for 24LPC or 26LPC.

## Future Shipments

The Newborn Screening Quality Assurance Program will ship next quarter's XALDPT specimens on July 10, 2018.

---

## Direct Inquiries

If you have any comments or questions about XALDPT MS/MS analysis, contact Dr. Christopher A. Haynes at 770-488-7019 or by e-mail at [cph7@cdc.gov](mailto:cph7@cdc.gov)

For data reporting questions, contact Irene Williams at [nsgapdmt@cdc.gov](mailto:nsgapdmt@cdc.gov)

The content of this report may also be located on our website at:  
[http://www.cdc.gov/labstandards/nsgap\\_reports.html](http://www.cdc.gov/labstandards/nsgap_reports.html)

*The identity of participants in any NSQAP proficiency testing scheme are considered confidential and known only to persons involved in the operation of the NSQAP proficiency testing scheme. Confidentiality may be waived by the participant upon written request only.*

**This program is co-sponsored by the Centers for Disease Control and Prevention (CDC) and  
The Association of Public Health Laboratories (APHL)**


## **NEWBORN SCREENING QUALITY ASSURANCE PROGRAM**

Direct inquiries to:

Centers for Disease Control and Prevention  
4770 Buford Highway NE, MS/F19  
Atlanta, GA 30341-3724  
Phone: 404-488-7945 Email: [cph7@cdc.gov](mailto:cph7@cdc.gov)

Editors

Chris Haynes  
Irene Williams



This *NEWBORN SCREENING QUALITY ASSURANCE PROGRAM* report is an internal publication distributed to program participants and selected program colleagues. The laboratory quality assurance program is a project cosponsored by the Centers for Disease Control and Prevention (CDC) and the Association of Public Health Laboratories.

**CENTERS FOR DISEASE CONTROL AND PREVENTION (CDC) ATLANTA, GA 30341**

**Acting Director**

Anne Schuchat, M.D. (RADM, USPHS)

**Director**

National Center for Environmental Health

Patrick Breysse, Ph.D.

**Director**

Division of Laboratory Sciences

James L. Pirkle, M.D., Ph.D.

**Chief**

Newborn Screening and Molecular Biology Branch

Carla Cuthbert, Ph.D.

**Contributors:**

Carter Asef

John Bernstein

Quan Bui

Preston Carter

Paul Dantonio

Sharon Flores

Elizabeth M. Hall

Christopher Haynes, Ph.D.

Brandon Kenwood

Francis Lee, Ph.D.

Lixia Li, Ph.D.

Timothy Lim, Ph.D.

Daniel Mandel, Ph.D.

Joanne Mei, Ph.D.

Kristina Mercer, Ph.D.

Gyliann Peña

Konstantinos Petritis, Ph.D.

Austin Pickens, Ph.D.

Robert Vogt, Ph.D.

Irene Williams

Sophia Winchester

Golriz Yazdanpanah

Sherri Zobel

**Production:**

Sarah Brown

Kimberly Coulter

Kizzy Stewart

**ASSOCIATION OF PUBLIC HEALTH LABORATORIES SILVER SPRING, MD 20910**

**President**

Ewa King, PhD

**Chairman, Newborn Screening and Genetics in Public Health Committee**

Michele Caggana, Sc.D., FACMG

**Chairman, Newborn Screening Quality Assurance Quality Control Subcommittee**

Patricia R. Hunt, B.A. and Joseph Orsini, Ph.D.

**Chairman, Newborn Screening Molecular Subcommittee**

Rachel Lee, Ph.D.

**INQUIRIES TO:**

Irene Williams, Editor • Centers for Disease Control and Prevention (CDC) • Newborn Screening Quality Assurance Program

Mailstop F-24 • 4770 Buford Highway, N.E. • Atlanta, GA 30341-3724

Phone (770) 488-4582 • NSQAPDMT@cdc.gov

E-mail: IWilliams1@cdc.gov