Newborn Screening Quality Assurance Program X-linked Adrenoleukodystrophy in Dried Blood Spots Proficiency Testing Program (XALDPT)

In co-sponsorship with Association of Public Health Laboratories (APHL) Provided by the Newborn Screening and Molecular Biology Branch Centers for Disease Control and Prevention 4770 Buford Highway NE, MS/F19 Atlanta, GA 30341-3724 Email: NSQAPDMT@cdc.gov

Quarterly Report Volume 4, No. 3 Issued: November 20, 2018

Report Authorization

This report has been reviewed and authorized by Dr. Joanne Mei, Laboratory Chief, Newborn Screening Quality Assurance Program.

Confidentiality Statement

NSQAP participant information and evaluations are strictly confidential and shared only with individual participants, unless written authorization for release is received.

Introduction

This report summarizes data collected within the specified period for Quarter 4, 2018, for the detection of X-ALD by analysis of the biomarkers 24:0-Lysophosphatidylcholine (24LPC) and 26:0-Lysophosphatidylcholine (26LPC) in dried blood spots (DBS). It is distributed to all participants, state laboratory directors, and program colleagues by request. The tables within this report provide certification profiles for the distributed specimens, statistical analysis of participant quantitative data, and frequency of clinical assessments. An evaluation of your laboratory's data is attached to this summary.

Certification of PT Specimens

DBS specimens were prepared from Type A+ human whole blood, which was adjusted to a hematocrit of $50 \pm 1\%$ and enriched with the biomarkers 24LPC and 26LPC. Expected values for each were determined by LC-MS/MS in units of µmol/L blood. Clinical assessments were based on the NSQAP cut-off of 0.47 µmol/L blood for 24LPC and 0.39 µmol/L blood for 26LPC. Table 1 shows the NSQAP expected values and clinical assessments for each specimen.

Table 1. Speciment Certification -	24I PC and 26I PC ((umol/L blood)
Table 1. Specifient Certification -	- 24LF C anu 20LF C (

Specimen	Expected 24LPC	24LPC Assessment Code*	Expected 26LPC	26LPC Assessment Code*
41821	0.07	1	0.04	1
41822	0.07	1	0.04	1
41823	0.07	1	0.04	1
41824	0.07	1	0.04	1
41825	1.07	2	1.04	2

*1 = Within Normal Limits

2 = Outside Normal Limits

Distribution of PT Specimens

On September 25, 2018 a PT panel of five unknown DBS specimens was distributed to 13 domestic laboratories and 5 foreign laboratories.

Participant Results

Quantitative Data

We processed data from 18 participants, with two participants submitting two method assessments. Laboratories were asked to report concentrations of 24LPC and 26LPC results in μ mol/L blood. Data not submitted in the requested units were not accepted. The conversion factor from μ g/mL to μ mol/L blood is provided on the XALDPT Data Report Form.

Overall statistics from MS/MS methods were combined so as to not identify an individual laboratory. We did not include data that were outside the 99% confidence interval. The statistical summary analysis for all methods is provided in Tables 2a-b.

Eight participants reported using Flow Injection Analysis (FIA) MS/MS non-kit, twelve reported using LC-MS/MS and two reported a two-tier assessment scheme utilizing both FIA– and LC-MS/MS. Seventeen participants submitted quantitative results for 24LPC, and four did not report a clinical assessment. Twenty participants reported quantitative results and clinical assessments for 26LPC. One participant reported cutoffs for 24LPC using multi-variate analysis by a post-analytic tool and 2nd tier testing when indicated. Tables 3a and 3b show the cutoffs reported for 24LPC and 26LPC sorted by method.

Analyte/Specimen	Ν	Mean (µmol/L)	SD
41821	17	0.14	0.08
41822	17	0.14	0.07
41823	17	0.14	0.07
41824	17	0.14	0.07
41825	17	1.14	0.29

Table 2a. Screening Results for 24LPC- All MS/MS Methods

Table 2b. Screening Results for 26LPC – All MS/MS Methods

Analyte/Specimen	Ν	Mean (µmol/L)	SD
41821	20	0.13	0.13
41822	20	0.13	0.13
41823	20	0.13	0.12
41824	20	0.13	0.13
41825	20	1.14	0.25

Table 3a. Analyte Cutoffs Sorted by Method (µmol/L)

Method 53 - LC-MS/MS	24LPC	26LPC
N	7	12
Mean	0.38	0.29
Max	0.55	0.50
Min	0.16	0.10
Median	0.39	0.25
Mode	NA	0.40

Table 3b. Analyte Cutoffs Sorted by Method (µmol/L)

Method 67 - FIA-MS/MS	24LPC	26LPC
N	5	7
Mean	0.48	0.40
Max	0.62	0.47
Min	0.33	0.36
Median	0.50	0.40
Mode	NA	0.36

Clinical Assessments

Laboratories were asked to report qualitative results as "Within Normal Limits" or "Outside Normal Limits". Qualitative assessments may differ because of specific assessment practices. The frequency distribution of participants' clinical assessments is shown in Tables 4a-b.

Table 4a. Frequency Distribution of Clinical Assessments for 24LPC*

Specimen	Within Normal Limits (WNL)	Outside Normal Limits (ONL)	No Assessment Reported for 24LPC
41821	13	0	7
41822	13	0	7
41823	13	0	7
41824	13	0	7
41825	0	13	7

*Two participants submitted clinical assessments for two methods (First- and Second-tier).

Table 4b. Frequency Distribution of Reported Clinical Assessments for 26LPC*

Specimen	Within Normal Limits (WNL)	Outside Normal Limits (ONL)
41821	20	0
41822	20	0
41823	20	0
41824	20	0
41825	0	20

*Two participant submitted clinical assessments for two methods (First- and Second-tier).

Evaluations

No False-negative and no False-positive misclassifications were reported for 24LPC. No False-negative and no False-positive misclassifications were reported for 26LPC.

Future Shipments

The Newborn Screening Quality Assurance Program will ship next quarter's PT specimens on January 15, 2019.

Direct Inquiries

If you have any comments or questions about XALDPT MS/MS analysis, contact Dr. Christopher A. Haynes at 770-488-7019 or by e-mail at cph7@cdc.gov

For data reporting questions, contact Irene Williams at nsqapdmt@cdc.gov

The content of this report may also be located on our website at: <u>https://www.cdc.gov/labstandards/nsqap_reports.html</u>

Acknowledgement

This *NEWBORN SCREENING QUALITY ASSURANCE PROGRAM* report is an internal publication distributed to program participants and selected program colleagues. The laboratory quality assurance program is a project cosponsored by the Centers for Disease Control and Prevention (CDC) and the Association of Public Health Laboratories.

CENTERS FOR DISEASE CONTROL AND PREVENTION (CDC) ATLANTA, GA 30341

Director

Robert R. Redfield, M.D.

Director

National Center for Environmental Health Patrick Breysse, Ph.D.

Director

Division of Laboratory Sciences James L. Pirkle, M.D., Ph.D.

Chief

Newborn Screening and Molecular Biology Branch Carla Cuthbert, Ph.D.

Contributors

- Carter AsefTimothy LiNicolle Baird, Ph.D.Daniel MarJohn BernsteinJoanne MarQuan BuiKristina MarPaul DantonioGyliann ParSharon FloresKonstantinElizabeth M. HallAustin PickChristopher Haynes, Ph.D.E. ShannoLaura Centeno HildrethRobert VogBrandon Kenwood, Ph.D.Irene WilliaFrancis Lee, Ph.D.Sophia WiLixia Li, Ph.D.Golriz YazSherri Zob
 - Timothy Lim, Ph.D. Daniel Mandel, Ph.D. Joanne Mei. Ph.D. Kristina Mercer, Ph.D. Gyliann Peña Konstantinos Petritis, Ph.D. Austin Pickens, Ph.D. E. Shannon Torres, Ph.D. Robert Vogt, Ph.D. Irene Williams Sophia Winchester Golriz Yazdanpanah Sherri Zobel

Production

Vinay Anamula Kimberly Coulter Joy Pressley Kizzy Stewart

ASSOCIATION OF PUBLIC HEALTH LABORATORIES SILVER SPRING, MD 20910

President

Joanne Bartkus, PhD, D(ABMM)

Chairman, Newborn Screening and Genetics in Public Health Committee Michele Caggana, Sc.D., FACMG

Chairman, Newborn Screening Quality Assurance Quality Control Subcommittee Patricia R. Hunt, B.A. and Joseph Orsini, Ph.D.

Chairman, Newborn Screening Molecular Subcommittee

Rachel Lee, Ph.D.

INQUIRIES TO:

Irene Williams, Editor Centers for Disease Control and Prevention (CDC), Newborn Screening Quality Assurance Program Mailstop F-24, 4770 Buford Highway, N.E., Atlanta, GA 30341-3724 E-mail: <u>NSQAPDMT@cdc.gov</u>