



Association for the  
Advancement of  
Blood & Biotherapies

# The AABB Experience with Addressing Common Nonconformances and How to Prevent Them



Clinical Laboratory Improvement Advisory Committee  
Wednesday, November 8, 2023

Melanie Sloan  
Senior Director, Accreditation and Quality  
AABB

## CMS TOP TEN DEFICIENCIES 1-5

Rank	Description	Regulatory Citation	D-Tag	Number of Times Cited	Percentage of Citations*
1	Personnel Competency Assessment (Standard-Level)	§ 493.1235	D5209	1,195	3.2%
2	General Lab Systems: Twice a year Accuracy Verification (Standard-Level)	§ 493.493.1236(c)(1)	D5217	1,057	2.9%
3	Analytic Systems: Reagent and specimen storage (Standard-Level)	§ 493.1252(b)	D5413	1,036	2.8%
4	Analytic Systems: Procedure manual requirements (Standard-Level)	§ 493.1251(b)	D5403	927	2.5%
5	Analytic Systems: Verification of Performance Specifications (Standard-Level)	§ 493.1253(b)(1)	D5421	775	2.1%

# Top 5 AABB Non-Conformances



<b>BBTS 2.1.3</b>	<b>5.6%</b>	<b>Evaluations of competence shall be performed before independent performance of assigned activities and at specified intervals.</b>
BBTS 1.3	5.6%	Quality and operational policies, processes, and procedures shall be developed and implemented to ensure that the requirements of these BB/TS Standards are satisfied. All such policies, processes, and procedures shall be in writing or captured electronically and shall be followed. Standard 5.1.1 applies.
CT 5.0	4.8%	The facility shall identify, design, modify, and validate the policies, processes, and procedures that affect the quality of cellular therapy products, services, and patient care.
BBTS 5.0	4.3%	The BB/TS shall have policies and validated processes and procedures that ensure the quality of the blood, blood components, tissue, derivatives, and services. The BB/TS shall ensure that these policies, processes, and procedures are carried out under controlled conditions.
BBTS 3.5	3.3%	The BB/TS shall have a process for scheduled monitoring and maintenance of equipment that at a minimum is in accordance with manufacturer's written instructions. The process shall include frequency of checks, check methods, acceptance criteria, and actions to be taken for unsatisfactory results.

# Examples of actual Nonconformance Statements for BBTS Standard 2.1.3

Standards 2.1.3: Evaluations of competence shall be performed before independent performance of assigned activities and at specified intervals.\*

\*[42 CFR 493.1235](#) and [42 CFR 493.1451\(b\)\(8\)\(9\)](#).

## Documented Nonconformance

Competency assessment does not include the requirement of Direct observation of performance of instrument maintenance and function checks.

Annual competency training for bedside nurse performing blood administration was also requested but not provided.

During an observation of a blood product administration on unit 4 East, nurse FH training records for blood administration was reviewed and it was found that she had not completed the annual requirement since 2020

No record of competency assessment for staff located for 2021 or 2022. The documents for 2020 were reviewed and acceptable. Facility is in transition with staff overturn and recent retirement of their blood bank manager, so the competency may have been done, however the records could not be located.

# Examples of actual Nonconformance Statements for BBTS Standard 1.3

Quality and operational policies, processes, and procedures shall be developed and implemented to ensure that the requirements of these BB/Ts Standards are satisfied. All such policies, processes, and procedures shall be in writing or captured electronically and shall be followed. Standard [5.1.1](#) applies.

## Documented Nonconformance

The Component Preparation Log SOP refers to acceptable product volumes for a 450ml bag and was not updated to reflect the acceptable volumes for the 500ml bag. The BDR is used to document volume of blood collected, however, the space used to record this is labeled grams instead of milliliters.

TR 615.6 Hemotemp II Indicators procedure does not follow manufacture's instruction for use

SOP # 1657: Polyethylene Glyco PEG Procedure 1.d. says "add two drops of patient's plasma to each testing tube" and Procedure 1.e. say "add the red cells to each testing tube" step-wise. Staff was observed and verbally confirmed of adding the red cells first and then adding the two drops of plasma to each of the testing tube in reverse order

# AABB Efforts to Mitigate these Common Nonconformances

[HOME](#) | [STANDARDS & ACCREDITATION](#) | [ABOUT AABB ACCREDITATION](#) | [AABB COMMENDABLE PRA](#)

## AABB COMMENDABLE PRACTICES

Resource Management

- [Agreement Contract Performance Review](#)

Improvement  
Management

Process Management

- [ABO and Rh Discrepancy Worksheet](#) (open

## AABB CLIA CORNER

Ask FDA/CMS

Competency  
Assessment

Test Complexity

Proficiency  
Testing

CLIA Certificates



# AABB Efforts to Mitigate these Common Nonconformances

HOME | STANDARDS & ACCREDITATION | ABOUT AABB ACCREDITATION | AABB COMMENDABLE PRA

## AABB COMMENDABLE PRACTICES

### Resource Management

- Agreement Contract Performance Review
- Assessment of Problem-Solving Skills
- Blood Bank Competency Checklist (open to non-members)



**Blood Bank Competency Checklist**

Name: \_\_\_\_\_ Laboratory Site: \_\_\_\_\_ Year \_\_\_\_\_  6-Month  Annual

Test System:	Element	Documentation Supplied-Medtech	Documentation Supplied-Pathnet	Date	Assessor Initials
Observe routine test performance, including, as applicable, patient identification and preparation, and specimen collection, handling, processing and testing.	1	Direct observation	Direct observation		
Reporting of Test Results	2	Internal Inquiry	Screen print of test result		
Review of QC records	3	Recorded on Direct Observation	Copy of QC Form		
Observe Instrument Maintenance (if applicable)	4	Recorded on Direct Observation	Copy of QC Form		
Assess Proficiency Testing or Blind sample	5	Report from original and test patient	Report from original and test patient		
Demonstrate Problem-Solving Skills	6	Problem solving case study	Problem solving case study		

I have had the opportunity to review and ask questions about policies and procedures related to equipment and testing above. Associate Signature: \_\_\_\_\_  
Date: \_\_\_\_\_

This associate is deemed to be competent to perform unsupervised patient testing in the above test systems. CLIA Medical Director/Designee Signature: \_\_\_\_\_  
Date: \_\_\_\_\_

## AABB CLIA CORNER

Ask FDA/CMS    Competency Assessment    Test Complexity    Proficiency Testing    CLIA Certificates

- CLIA proficiency testing and PT referral booklet
- Failed proficiency test report form
- PT report example
- CLIA Corner PT attestations
- **Twice Annual Verification nonconformances**



**AABB Standard 5.1.2 Proficiency Testing Program**  
The BIL/TS shall participate in a proficiency testing program, if available, for testing regulated by the Clinical Laboratory Improvement Amendments and performed by the facility. \* When a CMS-approved program is not available, there shall be a system for determining the accuracy and reliability of test results. Results shall be reviewed and when expected results are not achieved, investigation and corrective action shall be taken where appropriate.  
*\*42 CFR 493.1236.*

Per AABB, Chapter 5 Process Control is frequently cited. See the chart below for examples of the documented objective evidence that was found during assessments.

Standard	Objective Evidence
5.1.2	In 2019 only 1 eluate sample was present in an AABB IRL proficiency sample therefore twice annual verification was not performed.
5.1.2	The facility does have a process for a twice annual verification, but the assessor did not see any results
5.1.2	Proficiency testing for Microbial detection and flow cytometry CD34 testing has not been performed twice a year at a minimum in 2020. There were also gaps in 2019.

# AABB Efforts to Mitigate these Common Nonconformances

Participation in

## Workforce Action Alliance

Sponsored by:



Educational Offerings:

**Addressing Common Nonconformances from AABB: Deep Dive (23EL-894)**

*12/6/2023*

This program is a continuation of the conversation from the September 13th AABB eCast, Addressing Common Nonconformances from AABB and How to Prevent Them.

<https://www.aabb.org/education/calendar/event/2023/12/06/aabb-ecast/addressing-common-nonconformances-from-aabb-deep-dive-23el-894>

**Addressing Common Nonconformances from AABB and How to Prevent Them (21EL-694)**

*11/17/2021*

Open discussion with case studies to be presented by AABB's accreditation team. AABB standards and interpretations that resulted in nonconformances will be highlighted. The speakers will then expand discussion on selected items...

<https://www.aabb.org/education/calendar/event/2021/11/17/aabb-ecast/addressing-common-nonconformances-from-aabb-and-how-to-prevent-them-21el-694>







## Davidson Ballroom A2/A3

Tuesday, October 17

8:30 am - 10:00 am

[AM23-TU-06] A Quality Toolkit  
Conformance to AABB Standard

10:00 am - 11:00 am

[AM23-TU-11] Leadership and Personal Development:  
Tools to Unlearn Unconscious Bias (workshop)

SESSION IS FULL

# 2023 AABB Annual Meeting Session: A Quality Toolkit to Aid in Conformance to AABB Standard 5.0

The AABB Accreditation Staff will introduce the tools and use these quality concepts, and will dissect actual objective evidence cited for non-conformance to Standard 5.0:  
Process Control

## Current Quality Tools to Aid in Conformance

Process Improvement  
Check Sheet

Pareto Chart to Identify  
"Vital Few" Improvements

Root Cause Analysis  
Fishbone Diagram

Process Sequencing  
Flowchart

Process Pattern Analysis  
Histogram

Control Chart to  
Determine Statistical  
Variation

5 Whys to Identify a Root  
Cause and Solution

Affinity Diagram to  
Organize Ideas

Interrelationship Diagram  
to Identify Root Causes

Process Sequencing Tree  
Diagram

Regulatory Toolkits

Thank you!

