

# ***CDC Hormone Standardization Program (CDC HoSt)***

## ***Certified Total Testosterone Assays***

From 2019 Q4 and forward

(UPDATED 09/2023)

**CDC HoSt started quarterly certification from November 2019 shipment.**

See [https://www.cdc.gov/labstandards/csp/pdf/hs/CDC\\_Certified\\_Testosterone\\_Procedures-508.pdf](https://www.cdc.gov/labstandards/csp/pdf/hs/CDC_Certified_Testosterone_Procedures-508.pdf) for previous list of certified assays.

- The following assays have successfully met the performance criterion of  $\pm 6.4\%$  mean bias when compared to the CDC reference measurement procedure for total testosterone.
- It is not the intent of the CDC HoSt Program to certify each lot of reagents. Participants are awarded certificates for successfully meeting bias criterion using specific methods that consist of different reagent lots and calibrator lots.
- Analytical performance in CDC HoSt Program is assessed using human serum. The measurement procedures may have different accuracy and precision with other specimen types, such as plasma.
- Certification is valid for one quarter from the certification date. It is the responsibility of the participant to ensure that the results of the assay remain consistent, between lots, and over the measurement range reported.
- The analytical performance evaluation used in certification is for testing performed in patient care. Therefore, this certification does not imply suitability of a participant as a calibration laboratory or the procedure as a metrological reference measurement procedure.

Each table includes information about certified assays, including participant name, measurement principle, assay identifier, assay measurement range, certification measurement range, certification date, individual samples pass rate, and contact information.

“Assay identifier” is an internal code used by the participant to represent the assay used for certification.

“Assay Measurement range” is the assays’ reported analytical measurement range (AMR) and is not the certification range.

“Certification Measurement Range” is the concentration range the of the samples used for HoSt certification.

“Certification date” includes historical certification information and gaps between years do not always indicate the assay’s failure to meet certification criteria.

“Individual samples pass rate” is the percentage of individual samples out of the 40 provided that met the certification criteria of  $\pm 6.4\%$  bias. This information was provided starting in February 2017.

Note: The  $\pm 6.4\%$  mean bias used for certification can be considered the allowable calibration bias. Certification indicates that the assay is calibrated to meet those limits. Due to differences in test selectivity, measurements on individual samples can exceed the calibration bias. Therefore, the individual sample pass rate provides some information about the selectivity of a test that meets the calibration criteria.

CDC CSP suggests manufacturers and developers to participate in HoSt for certifications. The end-users or secondary location of an LDT may participate in Accuracy-based Monitoring Programs (CDC AMP) to verify performance. This document also indicates secondary LDT locations that are participating in AMP (Table 2).

**Table 1: Currently Certified Assays including their certification history**

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>ARUP Laboratories</b> Salt Lake City, UT	LC/MS/MS†	Total Testosterone in Serum	1.00 - 2,500	7.71 - 840	Q2 2023	62	Kayla West <a href="mailto:Kayla.west@aruplab.com">Kayla.west@aruplab.com</a> 801-583-2787 x2893
				7.71 - 840	Q1 2023	45	
				7.71 - 753	Q4 2022	68	
				7.71 - 753	Q3 2022	60	
				10.1 - 753	Q2 2022	42	
				10.1 - 736	Q1 2022	50	
				10.1 - 736	Q4 2021	38	
				8.53 - 680	Q3 2021	55	
				7.71 - 680	Q2 2021	75	
				7.71 - 680	Q1 2021	85	
				7.71 - 941	Q4 2020	90	
				7.71 - 941	Q3 2020	92	
				8.77 - 941	Q2 2020	72	
				9.63 - 941	Q1 2020	70	
5.70 - 840	Q4 2019	52					
<b>BioReference Laboratories, an OPKO Health Company</b> Elmwood Park, NJ	LC/MS/MS†	Total Testosterone	1 - 4,000	7.71 - 821	Q3 2020	62	Hashim Othman, Ph.D. <a href="mailto:hothman@bioreference.com">hothman@bioreference.com</a>
				8.49 - 915	Q2 2020	72	
				8.49 - 915	Q1 2020	68	
				8.49 - 915	Q4 2019	75	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Brigham Research Assay Core (BRAC) Laboratory at Harvard Medical School</b> Boston, MA	LC/MS/MS+	Serum Total Testosterone	1.00 - 2,000 (>2,000 with dilution)	9.51 - 840	Q2 2023	70	Dr. Shalender Bhasin <a href="mailto:sbhasin@bwh.harvard.edu">sbhasin@bwh.harvard.edu</a> (617) 525-9040  Liming Peng <a href="mailto:Lpeng2@bwh.harvard.edu">Lpeng2@bwh.harvard.edu</a> (617) 525-9048
				9.51 - 840	Q1 2023	65	
				10.9 - 840	Q4 2022	70	
				8.53 - 840	Q3 2022	78	
				7.71 - 753	Q2 2022	90	
				7.71 - 941	Q1 2022	92	
				7.71 - 941	Q4 2021	82	
				7.71 - 941	Q3 2021	72	
				8.77 - 941	Q2 2021	72	
				8.77 - 680	Q1 2021	68	
				8.77 - 941	Q4 2020	68	
				8.77 - 941	Q3 2020	57	
				8.77 - 941	Q2 2020	50	
				7.71 - 941	Q1 2020	55	
5.70 - 915	Q4 2019	57					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Clinical Chemistry Branch</b> <b>CDC</b> Atlanta, GA	LC/MS/MS <sup>+</sup>	Total Testosterone in Serum (1036)	0.57 - 12,800	7.71 - 680	Q1 2023	90	Lumi Duke, MS <a href="mailto:LDuke@cdc.gov">LDuke@cdc.gov</a> (770) 488-4126
				7.71 - 680	Q4 2022	90	
				7.71 - 629	Q3 2022	92	
				7.71 - 651	Q2 2022	88	
				8.53 - 651	Q1 2022	90	
				8.53 - 680	Q4 2021	90	
				8.53 - 840	Q3 2021	90	
				11.0 - 840	Q2 2021	90	
				8.77 - 840	Q1 2021	92	
				8.77 - 941	Q4 2020	95	
				8.49 - 941	Q3 2020	92	
				8.49 - 941	Q2 2020	98	
				8.49 - 941	Q1 2020	98	
8.49 - 915	Q4 2019	98					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Covance Central Laboratories Services, Inc.</b> Indianapolis, IN	LC/MS/MS+	Serum Total Testosterone	2.00 - 8,000	9.51 - 941	Q2 2023	80	Cristina Hedin, MS 8211 Scicor Drive Indianapolis, IN 46214 <a href="mailto:cristina.hedin@covance.com">cristina.hedin@covance.com</a> (317) 273-7842
				9.51 - 941	Q1 2023	90	
				9.51 - 941	Q4 2022	87	
				9.51 - 941	Q3 2022	79	
				8.77 - 941	Q2 2022	82	
				8.77 - 769	Q1 2022	82	
				7.71 - 840	Q4 2021	78	
				7.71 - 840	Q3 2021	88	
				7.71 - 840	Q2 2021	78	
				7.71 - 840	Q1 2021	78	
				8.77 - 941	Q4 2020	82	
				8.77 - 941	Q3 2020	78	
				8.77 - 941	Q2 2020	85	
				7.71 - 941	Q1 2020	85	
7.71 - 753	Q4 2019	80					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Covance Central Laboratories Services, Inc.</b> Geneva, Switzerland	LC/MS/MS <sup>+</sup>	Total Testosterone	2.00 - 8,000	13.4 - 821	Q2 2023	85	Gaelle Gilbert <a href="mailto:gaelle.gilbert@covance.com">gaelle.gilbert@covance.com</a> 0041 (0) 58 822 7656
				12.0 - 821	Q1 2023	88	
				12.0 - 629	Q4 2022	88	
				12.0 - 941	Q3 2022	95	
				12.0 - 941	Q2 2022	98	
				12.2 - 941	Q1 2022	90	
				12.2 - 941	Q4 2021	95	
				12.2 - 769	Q3 2021	75	
				12.2 - 769	Q2 2021	70	
				8.77 - 821	Q1 2021	75	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Diagnostic Laboratory for Endocrinology, Erasmus University Medical Center</b>  Rotterdam, The Netherlands	LC/MS/MS+	Serum Testosterone	2 - 1,093	9.51 - 769	Q2 2023	85	S.A.A. van den Berg <a href="mailto:s.a.a.vandenberg@erasmusmc.nl">s.a.a.vandenberg@erasmusmc.nl</a> S.S. Panchoe - Ramcharan <a href="mailto:s.panchoe@erasmusmc.nl">s.panchoe@erasmusmc.nl</a>
				7.71 - 821	Q1 2023	85	
				7.71 - 821	Q4 2022	90	
				7.71 - 941	Q3 2022	88	
				7.71 - 941	Q2 2022	90	
				8.77 - 941	Q1 2022	92	
				7.71 - 941	Q4 2021	80	
				7.71 - 769	Q3 2021	85	
				7.71 - 769	Q2 2021	70	
				7.71 - 736	Q1 2021	68	
				8.77 - 753	Q4 2020	78	
				8.49 - 821	Q3 2020	78	
				8.49 - 821	Q2 2020	92	
				8.49 - 840	Q1 2020	92	
8.49 - 840	Q4 2019	95					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Endocrine and Metabolic Research Laboratory at Los Angeles Biomedical Research Institute</b> Torrance, CA	LC/MS/MS <sup>+</sup>	TDHT	2.0 - 2,000	9.51 - 941	Q2 2023	55	Dr. Christina Wang <a href="mailto:wang@lundquist.org">wang@lundquist.org</a> (310) 222-2503
				7.71 - 941	A1 2023	50	
				7.71 - 941	Q4 2022	50	
				11.0 - 769	Q1 2022	50	
				8.53 - 753	Q4 2021	52	
				8.53 - 753	Q3 2021	65	
				8.53 - 736	Q2 2021	68	
				8.53 - 736	Q1 2021	50	
				8.77 - 941	Q4 2020	52	
				8.77 - 941	Q2 2020	48	
3.66 - 941	Q1 2020	70					
<b>GC Labs</b> Yongin-si, Gyeonggi-do, Republic of Korea	LC/MS/MS <sup>+</sup>	Total Testosterone	0.8 - 5188.5	8.53 - 941	Q2 2023	92	Sung-Eun Cho, M.D., Ph.D <a href="mailto:secho1206!!@gclabs.co.kr">secho1206!!@gclabs.co.kr</a> Jung-Sun Han <a href="mailto:jshan@gclabs.co.kr">jshan@gclabs.co.kr</a>
				8.53 - 941	Q1 2023	92	
				8.53 - 821	Q4 2022	92	
				8.77 - 941	Q3 2022	95	
				8.53 - 736	Q2 2022	88	



Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>LabCorp</b> Calabasas Hills, CA	LC/MS/MS+	#070001 Testosterone, Total, Women, Children, and Hypogonadal Males, LC MS/MS	2.50 - 5,000  (250,000 with validated dilution)	7.71 - 840	Q2 2023	85	Majid Moridani <a href="mailto:moridam@labcorp.com">moridam@labcorp.com</a> 336-436-3102  Dr. Andre Valcour <a href="mailto:ValcouA@labcorp.com">ValcouA@labcorp.com</a> (336) 436-3854  Dr. Brett Holmquist <a href="mailto:holmqub@labcorp.com">holmqub@labcorp.com</a> (818) 867-1362
				7.71 - 840	Q1 2023	80	
				8.77 - 840	Q4 2022	72	
				10.9 - 753	Q3 2022	72	
				8.77 - 753	Q2 2022	72	
				8.77 - 941	Q1 2022	80	
				8.77 - 941	Q4 2021	95	
				8.77 - 941	Q3 2021	90	
				9.51 - 941	Q2 2021	88	
				8.77 - 941	Q4 2020	78	
				8.77 - 941	Q3 2020	82	
				8.77 - 941	Q2 2020	82	
				13.1 - 941	Q1 2020	88	
9.63 - 840	Q4 2019	90					
<b>LabCorp</b> Burlington, NC	LC/MS/MS+	#070001 Testosterone, Total, Women, Children, and Hypogonadal Males, LC MS/MS	2.50 - 5,000  (250,000 with validated dilution)	8.53 - 941	Q2 2021	90	Majid Moridani <a href="mailto:moridam@labcorp.com">moridam@labcorp.com</a> 336-436-3102
				8.77 - 941	Q1 2021	95	
				8.77 - 941	Q4 2020	95	
				8.77 - 941	Q3 2020	92	
				8.77 - 941	Q2 2020	95	
				8.49 - 941	Q1 2020	98	
				8.49 - 840	Q4 2019	98	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>LabCorp</b> Spokane, WA	LC/MS/MS+	Total Testosterone	2.5 – 1,000 (250,000 with validated dilution)	9.51 - 941	Q2 2023	82	Carissa Schmitz MLS(ASCP)CM <a href="mailto:Schmic4@LabCorp.com">Schmic4@LabCorp.com</a> (509) 755-8358
				9.51 - 941	Q1 2023	68	
				10.1 - 840	Q4 2022	68	
				7.71 - 840	Q3 2022	52	
				7.71 - 840	Q2 2022	48	
				7.71 - 736	Q1 2022	55	
				7.71 - 753	Q4 2021	57	
				8.77 - 753	Q3 2021	65	
				9.51 - 753	Q2 2021	62	
				8.77 - 821	Q1 2021	62	
				8.77 - 941	Q4 2020	70	
				8.49 - 941	Q3 2020	72	
				8.49 - 941	Q2 2020	80	
7.71 - 941	Q1 2020	88					
7.71 - 915	Q4 2019	88					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Mayo Clinic</b> Rochester, MN	LC/MS/MS+	Total Testosterone	7 – 2,000	8.53 - 821	Q2 2023	75	Sue Reicks <a href="mailto:reicks.sue@mayo.edu">reicks.sue@mayo.edu</a>
				8.53 - 821	Q1 2023	65	
				8.53 - 821	Q4 2022	65	
				8.53 - 840	Q3 2022	62	
				9.51 - 840	Q2 2022	55	
				9.51 - 941	Q1 2022	70	
				10.1 - 941	Q4 2021	60	
				10.1 - 941	Q3 2021	62	
				10.1 - 941	Q2 2021	62	
				8.77 - 651	Q1 2021	55	
				8.77 - 941	Q4 2020	57	
				8.77 - 941	Q3 2020	60	
				8.77 - 941	Q2 2020	72	
10.6 - 941	Q1 2020	70					
9.63 - 915	Q4 2019	72					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Penn State University Hershey Medical Center</b> Hershey, PA	LC/MS/MS <sup>+</sup>	Total Testosterone in Serum	2 - 2,330	7.71 - 736	Q2 2023	65	Yusheng Zhu, PhD, DABCC, FAACC <a href="mailto:yzhu@pennstatehealth.psu.edu">yzhu@pennstatehealth.psu.edu</a> (717) 531-5123
				7.71 - 651	Q1 2023	70	
				8.53 - 736	Q4 2022	68	
				8.53 - 736	Q3 2022	57	
				8.53 - 736	Q2 2022	60	
				8.77 - 941	Q1 2022	60	
				7.71 - 941	Q4 2021	65	
				7.71 - 941	Q3 2021	72	
				7.71 - 941	Q2 2021	70	
				7.71 - 753	Q1 2021	68	
				8.77 - 821	Q4 2020	68	
				8.77 - 821	Q3 2020	68	
				8.77 - 821	Q2 2020	75	
				9.63 - 840	Q1 2020	78	
9.63 - 941	Q4 2019	82					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Quest Diagnostics Nichols Institute of Valencia, Inc.</b> Valencia, CA  See <b>Table 2</b> for AMP status	LC/MS/MS+	Serum Total Testosterone	2 - 2,000 (10,000 with dilution)	8.53 - 680 7.71 - 680 7.71 - 680 7.71 - 680 7.71 - 680 8.77 - 680 8.77 - 769 8.77 - 821 11.0 - 941 8.77 - 941 8.77 - 941 8.49 - 941 8.49 - 821 8.49 - 821 8.49 - 941	Q2 2023 Q1 2023 Q4 2022 Q3 2022 Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	90 88 85 80 80 78 72 75 65 62 65 65 72 70 70	Amit Ghoshal PhD <a href="mailto:Amit.K.Ghoshal@QuestDiagnostics.com">Amit.K.Ghoshal@QuestDiagnostics.com</a> (661) 799-6204
<b>Roche Diagnostics GmbH</b> Penzberg, Germany	LC/MS/MS+	Total Testosterone in Serum	0.8 – 1,800	8.53 - 840 8.53 - 821 8.53 - 821 9.51 - 769 10.1 - 769 8.77 - 941 8.77 - 941	Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021 Q4 2020	80 68 75 65 70 70 70	Rupert Schmid <a href="mailto:rupert.schmid@roche.com">rupert.schmid@roche.com</a> +498856-605256

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>Siemens Healthcare Diagnostics</b> Newark, DE	Chemiluminescence Immunoassay	Serum Total Testosterone	8 - 1,000	11.0 - 941	Q4 2022	28	Dr. Craig Hixson <a href="mailto:craig.hixson@siemens.com">craig.hixson@siemens.com</a> (302) 631-7540
				11.0 - 941	Q3 2022	33	
				11.0 - 736	Q2 2022	31	
				10.9 - 680	Q1 2022	35	
				10.9 - 680	Q4 2021	31	
				9.51 - 680	Q3 2021	21	
				8.77 - 941	Q4 2020	24	
				8.49 - 941	Q3 2020	30	
				8.49 - 941	Q2 2020	24	
				8.49 - 941	Q1 2020	26	
8.49 - 915	Q4 2019	32					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Siemens Healthcare Diagnostics Tarrytown, NY	Chemiluminescence Immunoassay	ADVIA Centaur® Testosterone II Assay	7.0 – 1,500	10.9 - 941	Q2 2023	25	Neil Parker <a href="mailto:neil.np.parker@siemens-healthineers.com">neil.np.parker@siemens-healthineers.com</a> (914) 524-2477
				10.9 - 941	Q1 2023	25	
				10.9 - 941	Q4 2022	20	
				10.9 - 941	Q3 2022	22	
				10.1 - 769	Q2 2022	22	
				10.1 - 840	Q1 2022	22	
				10.1 - 840	Q4 2021	32	
				8.53 - 840	Q3 2021	28	
				10.9 - 840	Q2 2021	35	
				8.77 - 821	Q1 2021	35	
				8.77 - 941	Q4 2020	38	
				8.77 - 941	Q3 2020	38	
				8.77 - 941	Q2 2020	38	
				7.71 - 941	Q1 2020	35	
7.71 - 840	Q4 2019	32					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date <i>(active for 1 quarter)</i>	Individual Samples Pass Rate (%)	Participant's Contact Information
Siemens Healthcare Diagnostics Tarrytown, NY	Chemiluminescence Immunoassay	Atellica® Testosterone	7.0 – 1,500	8.53 - 769	Q2 2023	30	Neil Parker <a href="mailto:neil.np.parker@siemens-healthineers.com">neil.np.parker@siemens-healthineers.com</a> (914) 524-2477
				8.53 - 769	Q1 2023	28	
				8.53 - 769	Q4 2022	25	
				8.53 - 840	Q3 2022	22	
				8.53 - 941	Q2 2022	28	
				8.53 - 941	Q1 2022	26	
				8.53 - 941	Q4 2021	26	
				8.53 - 941	Q3 2021	31	
				10.1 - 736	Q2 2021	21	
				8.77 - 659	Q1 2021	18	
				8.77 - 821	Q4 2020	18	
				8.49 - 821	Q3 2020	15	
				8.49 - 821	Q2 2020	20	
				8.49 - 821	Q1 2020	28	
8.49 - 941	Q4 2019	32					



Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (ng/dL)	Certification Measurement Range (ng/dL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Siemens Healthcare Diagnostics Tarrytown, NY	Chemiluminescence Immunoassay	ADVIA Centaur CP	7.0 – 1,500	9.51 - 840	Q2 2023	25	Neil Parker <a href="mailto:neil.np.parker@siemens-healthineers.com">neil.np.parker@siemens-healthineers.com</a> (914) 524-2477
				10.1 - 840	Q1 2023	20	
				10.1 - 840	Q4 2022	25	
				10.1 - 840	Q3 2022	30	
				10.1 - 769	Q2 2022	31	
				9.51 - 769	Q1 2022	33	
				9.51 - 736	Q4 2021	28	
				10.1 - 840	Q3 2021	35	
				8.53 - 840	Q2 2021	38	
				8.53 - 840	Q1 2021	28	
				8.53 - 941	Q4 2020	35	
				8.49 - 941	Q3 2020	32	
University of Minnesota (MEBRL) Minneapolis, MN	LC/MS/MS†	Total Testosterone in Serum	2 – 2,000	8.53 - 753	Q4 2021	80	Revati Koratkar <a href="mailto:kora0033@umn.edu">kora0033@umn.edu</a> 612-624-2959
				7.71 - 753	Q3 2021	70	
				7.71 - 753	Q2 2021	60	
				7.71 - 659	Q1 2021	57	
				8.49 - 941	Q2 2020	75	

† LC/MS/MS – Liquid Chromatography Tandem Mass Spectrometry

‡ GC/MS/MS – Gas Chromatography Tandem Mass Spectrometry

**Table 2: Accuracy-Based Monitoring Programs (AMP) status of secondary location**

Participant	Measurement Principle	Assay Identifier	AMP Active Date <i>(active for 1 quarter)</i>	Participant's Contact Information
<p><b>Quest Diagnostics</b> Chantilly, VA</p>	<p>LC/MS/MS†</p>	<p>Serum Total Testosterone</p>	<p>Q2 2023 Q1 2023 Q4 2022 Q3 2022 Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019</p>	<p>William Wu PhD <a href="mailto:William.W.Wu@QuestDiagnostics.com">William.W.Wu@QuestDiagnostics.com</a> (703) 802-7210</p>