

CDC Vitamin D Standardization-Certification Program (CDC VDSCP)

Certified Total 25-hydroxyvitamin D Assays

From 2019 Q4 and forward

(UPDATED 11/2023)

CDC VDSCP started quarterly certification from November 2019 shipment.

See https://www.cdc.gov/labstandards/csp/pdf/hs/CDC_Certified_Vitamin_D_Procedures-508.pdf for previous list of certified assays.

- The following assays have successfully met the performance criterion of $\pm 5.0\%$ mean bias when compared to the CDC reference measurement procedure and an overall imprecision of $<10\%$. **Only data for total 25-hydroxyvitamin D was used for certification.**
- It is not the intent of the CDC VDSCP to certify each lot of reagents. Participants are awarded certificates for successfully meeting bias and imprecision criteria using specific methods that consist of different reagent and calibrator lots.
- Analytical performance in the CDC VDSCP is assessed using human serum. The measurement procedures may have different accuracy and precision with other specimen types, such as plasma.
- Certification is valid for one quarter from the certification date. It is the responsibility of the participant to ensure that the results of the assay remain consistent, between lots, and over the measurement range reported.
- The analytical performance evaluation used in certification is for testing performed in patient care. Therefore, this certification does not imply suitability of a participant as a calibration laboratory or the procedure as a metrological reference measurement procedure.

Each table includes information about certified assays including participant name, measurement principle, assay identifier, certification measurement range, certification date, individual samples pass rate, and contact information.

“Assay identifier” is an internal code used by the participant to represent the assay used for certification.

“Assay Measurement range” is the assay’s reported analytical measurement range (AMR) and is not the certification range.

“Certification Measurement Range” is the concentration range the of the samples used for VDSCP certification.

“Certification date” includes historical certification information and gaps between years do not always indicate the assay’s failure to meet certification criteria.

“Individual samples pass rate” is the percentage of individual samples out of the 40 provided that met certification criteria of $\pm 5\%$ bias. This information was provided starting February 2017.

Note: The $\pm 5\%$ mean bias used for certification can be considered the allowable calibration bias. Certification indicates that the assay is calibrated to meet those limits. Due to differences in test selectivity, measurements on individual samples can exceed the calibration bias. Therefore, the individual sample pass rate provides some information about the selectivity of a test that meets the calibration criteria.

Table 1: Currently Certified Assays

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Abbott GmbH Longford, Ireland	Chemiluminescence Immunoassay	ALINITY 25-OH Vitamin D (List Number 08P45)	8.5 - 385.5	17.4 - 149	Q2 2021	45	Ronan O'Neill Ronan.O'Neill@abbott.com +353 (0)43-33-31008
				16.7 - 149	Q1 2021	42	
				13.1 - 183	Q1 2020	25	
				13.1 - 183	Q4 2019	32	
Abbott GmbH Longford, Ireland	Chemiluminescence Immunoassay	ARCHITECT 25-OH Vitamin D (List Number 5P02)	8.5 - 389.8	16.1 - 208	Q3 2023	33	Ronan O'Neill Ronan.O'Neill@abbott.com +353 (0)43-33-31008
				16.1 - 123	Q2 2021	40	
				13.1 - 136	Q1 2021	38	
				13.1 - 149	Q4 2020	38	
				12.1 - 220	Q3 2020	35	
				12.1 - 220	Q2 2020	28	
				12.1 - 220	Q1 2020	28	
12.1 - 220	Q4 2019	35					
Affimedix, Inc. Hayward, CA	Sandwich ELISA	MicrO - D™ 25-OH Vitamin D (total)	4.2 - 375	16.1 - 175	Q4 2019	35	Kevin Wang Kwang@affimedix.com

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
ARUP Laboratories, Inc. Salt Lake City, UT	LC/MS/MS+	25-Hydroxyvitamin D2 and D3 by Tandem Mass Spectrometry, Serum	2.5 - 249.6	12.1 - 157	Q1 2023	50	Kayla West Kayla.west@aruplab.com 801-583-2787 x2893
				12.1 - 142	Q1 2023	50	
				12.1 - 139	Q4 2022	65	
				12.1 - 162	Q3 2022	65	
				16.1 - 305	Q2 2022	48	
				16.1 - 305	Q1 2022	55	
				16.1 - 305	Q4 2021	60	
				16.7 - 305	Q3 2021	50	
				16.7 - 135	Q2 2021	65	
				16.1 - 135	Q1 2021	62	
				16.1 - 149	Q4 2020	65	
				16.1 - 149	Q3 2020	78	
				16.1 - 149	Q2 2020	65	
				18.1 - 183	Q1 2020	65	
19.7 - 183	Q4 2019	60					
AUTOBIO DIAGNOSTICS CO., LTD. Zhengzhou, China	Chemiluminescence Immunoassay	AutoLumo 25-OH Vitamin D Total Assay	11.25 - 375	13.1 - 140	Q2 2020	26	Zhang Yujiao zhangyujiao@autobio.com.cn 86-371-66785303

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Boditech Med Inc. South Korea	Fluorescence Immunoassay	Boditech Med Ichroma Vitamin D 25(OH)D2/D3	20 - 175	22.0 - 158	Q3 2023	20	Jieun Lee lje@boditech.co.kr 82-70-4264-7575
				22.6 - 157	Q2 2023	22	
				22.0 - 157	Q1 2023	21	
				22.0 - 162	Q3 2022	29	
				22.0 - 162	Q2 2022	32	
				22.6 - 139	Q3 2021	32	
				22.6 - 123	Q2 2021	42	
				22.6 - 183	Q1 2021	38	
				22.0 - 183	Q4 2020	38	
				12.1 - 183	Q3 2020	36	
				12.1 - 183	Q2 2020	28	
				12.1 - 183	Q1 2020	25	
12.1 - 183	Q4 2019	22					
Boditech Med Inc. South Korea	Fluorescence Immunoassay	Boditech Med AFIAS Vitamin D 25(OH)D2/D3	20 - 175	22.0 - 157	Q3 2023	11	Jieun Lee lje@boditech.co.kr 82-70-4264-7575
				22.0 - 162	Q2 2023	11	
				22.0 - 162	Q1 2023	8	
				22.0 - 162	Q3 2022	10	
				28.5 - 157	Q3 2021	40	
				28.5 - 140	Q2 2021	38	
				12.1 - 140	Q1 2021	31	
				12.1 - 135	Q4 2020	31	
				12.1 - 140	Q3 2020	24	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
CCR Laboratory University of Maryland Clinical Pathology Baltimore, MD	LC/MS/MS ⁺	Total 25 OH-Vitamin D 10 – 400 nmol/mL	10 - 400	12.1 - 208	Q3 2023	50	Dr. Show-Hong Duh sduh@umm.edu (410) 328-5705
				12.1 - 208	Q2 2023	38	
				12.1 - 208	Q1 2023	48	
				22.0 - 162	Q4 2022	38	
				12.1 - 162	Q3 2022	38	
				12.1 - 162	Q2 2022	42	
				12.1 - 162	Q1 2022	32	
				12.1 - 162	Q4 2021	38	
				16.7 - 158	Q3 2021	42	
				17.4 - 140	Q2 2021	50	
				16.7 - 183	Q1 2021	50	
				13.1 - 183	Q3 2020	38	
				13.1 - 183	Q2 2020	35	
				13.1 - 149	Q1 2020	42	
13.1 - 136	Q4 2019	52					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Covance Central Laboratory Services, Inc. Indianapolis, IN	LC/MS/MS†	Total 25 Hydroxy-Vitamin D2/D3	1.25 - 1,996.80	16.1 - 208	Q1 2022	92	Cristina Hedin cristina.hedin@covance.com (317) 273-7842
				16.1 - 208	Q4 2021	95	
				16.1 - 157	Q3 2021	80	
				16.1 - 135	Q2 2021	75	
				12.1 - 183	Q1 2021	72	
				12.1 - 183	Q4 2020	62	
				12.1 - 183	Q3 2020	78	
				12.1 - 183	Q2 2020	82	
				16.1 - 140	Q1 2020	82	
				18.1 - 140	Q4 2019	90	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Centre Hospitalier Universitaire de Liège Sart-Tilman, Belgium	LC/MS/MS†	25(OH)D3+25(OH)D2+C3-epimer+24,25(OH)2D	3.7 - 624	17.4 - 182 16.7 - 182 16.7 - 182 16.7 - 182 16.7 - 157 18.1 - 162 18.1 - 162 21.5 - 162 21.5 - 162 21.5 - 128 12.1 - 136 12.1 - 149 12.1 - 149 12.1 - 198 18.1 - 198 18.1 - 198	Q3 2023 Q2 2023 Q1 2023 Q4 2022 Q3 2022 Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	72 70 62 70 72 70 68 70 80 85 92 92 85 75 62 57	Prof. Etienne Cavalier etienne.cavalier@chu.ulg.ac.be +32 4 3668822
Department of Obstetrics and Gynaecology, The University of Hong Kong Hong Kong, China	Immunoassay	ARCHITECT® 25-OH Vitamin D Assay	8.5 - 389.8	17.4 - 208 17.4 - 182 16.1 - 158	Q3 2022 Q2 2022 Q4 2021	38 40 38	Raymond Li raymondli@hku.hk 85222554517

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
DiaSorin Stillwater, MN	Chemiluminescence Immunoassay	LIAISON® 25 OH Vitamin D TOTAL Assay	10 - 375	16.7 - 157	Q1 2023	22	Cole Gilman cole.gilman@diasorin.com (651) 351-5933
				17.4 - 157	Q4 2022	20	
				17.4 - 162	Q3 2022	20	
				17.4 - 305	Q2 2022	22	
				16.7 - 305	Q1 2022	30	
				16.7 - 305	Q4 2021	30	
				16.1 - 305	Q3 2021	30	
				13.1 - 183	Q2 2021	22	
				13.1 - 183	Q1 2021	12	
				13.1 - 183	Q4 2020	10	
				13.1 - 138	Q3 2020	10	
				16.7 - 220	Q2 2020	12	
				17.4 - 220	Q1 2020	15	
17.4 - 220	Q4 2019	28					
Diazyme Laboratories, Inc. Poway, CA	Immunoturbidimetric Assay	Diazyme EZ Vitamin D Total Assay	18.9 - 369	32.4 - 182	Q4 2022	12	Dr. Chao Dou chao.dou@diazyme.com 858-455-4763
				27.5 - 194	Q2 2022	13	
				22.2 - 194	Q1 2022	16	
				21.5 - 194	Q4 2021	8	
				21.5 - 194	Q3 2021	8	
				21.5 - 140	Q2 2021	14	
				16.7 - 140	Q1 2021	11	
				16.7 - 149	Q2 2020	19	
				18.1 - 149	Q1 2020	21	
				19.7 - 136	Q4 2019	18	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Douglass Hanly Moir Pathology Sydney, Australia	LC/MS/MS+	25-Hydroxyvitamin D2 and D3 by Tandem Mass Spectrometry, Serum	3.90 - 600	12.1 - 162 16.7 - 194 16.7 - 194 16.7 - 140 13.1 - 183 12.1 - 183	Q1 2023 Q4 2022 Q1 2022 Q4 2021 Q2 2021 Q1 2021 Q4 2020	10 32 45 55 55 68	Andy Liu aliu@dhm.com.au +61298555275
Endocrine Sciences Laboratory, LabCorp Calabasas Hills, CA	LC/MS/MS+	Test Code 500510 (Total)	2.4 - 625	12.1 - 208 12.1 - 208 12.1 - 208 12.1 - 208 12.1 - 149 17.4 - 208 16.1 - 208 16.1 - 208 16.1 - 208 13.1 - 140 13.1 - 149 13.1 - 149 10.4 - 149 10.4 - 175 10.4 - 175 10.4 - 183	Q3 2023 Q2 2023 Q1 2023 Q4 2022 Q3 2022 Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	70 68 62 60 60 60 65 70 78 72 68 62 57 65 72 70	Dr. Brett Holmquist Holmqub@labcorp.com (818)867-1362

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Fujirebio INC. Tokyo, Japan	Chemiluminescence Enzyme Immunoassay	Lumipulse® G 25-OH Vitamin D	10 - 375	16.1 - 183	Q1 2020	45	Satoshi Saito satoshi.kojima@miraca.com +81-42-645-4742
GC Labs Yongin-si, Gyeonggi-do, Republic of Korea	LC/MS/MS†	Serum Total 25OHD, 25OHD3, 25OHD2 and 3- epi-25OHD3	D3: 0.871- 364 D2: 1.21 - 329 epi: 1.25 - 67	12.1 - 208 12.1 - 208 12.1 - 208 17.4 - 162 17.4 - 149 17.4 - 182	Q3 2023 Q2 2023 Q1 2023 Q4 2022 Q3 2022 Q2 2022	85 82 75 70 68 68	Sung-Eun Cho, M.D., Ph.D. secho1206!!@gclabs.co.kr Jung-Sun Han jshan@gclabs.co.kr
Heartland Assays Ames, IA	LC/MS/MS†	25(OH)D2/D3 via LC/MS/MS	3.75 - 500	17.4 - 158 12.1 - 158 12.1 - 162 18.1 - 211 22.2 - 211 22.2 - 211 17.4 - 211 17.4 - 132 16.7 - 136 12.1 - 136 12.1 - 136 12.1 - 183 12.1 - 183 13.1 - 183	Q3 2023 Q2 2023 Q1 2023 Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	60 50 62 52 48 70 75 67 65 57 65 75 65 68	Drew Makowski Drew.makowski@heartlandassays.com or John Rathmacher John.rathmacher@heartlandassays.com

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Immunodiagnostic Systems (IDS) Tyne and Wear, GB	ELISA	25-Hydroxy Vitamin D EIA	16.3 - 250	16.1 - 208	Q2 2023	35	Heather Pham Heather.pham@idsplc.com
				16.1 - 157	Q1 2023	33	
				16.1 - 157	Q4 2022	33	
				17.4 - 182	Q2 2022	26	
				16.7 - 182	Q1 2022	20	
				16.7 - 182	Q4 2021	18	
				16.7 - 182	Q3 2021	26	
				16.7 - 134	Q2 2021	23	
				16.1 - 220	Q1 2021	29	
				16.1 - 220	Q4 2020	35	
				16.1 - 220	Q3 2020	24	
				16.1 - 220	Q2 2020	29	
				10.4 - 183	Q1 2020	24	
				10.4 - 183	Q4 2019	18	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Immunodiagnostic Systems (IDS) Tyne and Wear, GB	CLIA	IDS 25 VitD ^S	10 - 275	12.1 - 162	Q2 2023	45	Heather Pham Heather.pham@idsplc.com
				12.1 - 162	Q1 2023	42	
				12.1 - 158	Q4 2022	35	
				12.1 - 149	Q3 2022	28	
				17.4 - 162	Q2 2022	25	
				12.1 - 162	Q1 2022	22	
				12.1 - 162	Q4 2021	30	
				12.1 - 162	Q3 2021	38	
				12.1 - 135	Q2 2021	40	
				16.1 - 136	Q1 2021	45	
				16.1 - 136	Q4 2020	38	
				12.1 - 138	Q3 2020	33	
				10.4 - 183	Q1 2020	24	
10.4 - 183	Q4 2019	26					
Instituto Nacional de Ciencias Medicas Salvador Zubirán Mexico City, Mexico	CLIA	ARCHITECT® 25-OH Vitamin D Assay	8.5 – 389.8	16.1 - 182	Q4 2022	35	QFB. Alejandra Pacheco Cervantes alepachecoc@hotmail.com
Mayo Clinic Florida Jacksonville, FL	LC/MS/MS†	25 hydroxyvitamin D2 and D3	D2: 5 - 250 D3: 10 - 250	12.1 - 208 12.1 - 208	Q3 2023 Q2 2023	55 55	Charles Barmore barmore.charles@mayo.edu (904) 953-6411

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Mayo Medical Laboratories Rochester, MN	LC/MS/MS ⁺	Serum 25 hydroxyvitamin D2 and D3	15 - 740	17.4 - 162	Q1 2022	55	Larry Dodge Dodge.larry@mayo.edu (507) 538-0400 Eric Bro Bro.eric@mayo.edu (507) 266-4348
				17.4 - 162	Q4 2021	55	
				17.4 - 157	Q3 2021	50	
				17.4 - 135	Q2 2021	50	
				18.1 - 135	Q1 2021	52	
				16.1 - 149	Q4 2020	55	
				10.4 - 183	Q3 2020	62	
				10.4 - 183	Q2 2020	62	
				10.4 - 183	Q1 2020	67	
				10.4 - 183	Q4 2019	62	

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Nutrition Research Division Health Canada Ottawa, Canada	LC-MS/MS ⁺	Serum Total 25OHD, 25OHD3, 25OHD2 and 3-epi-25OHD3	2.5 - 275	16.1 - 208	Q2 2023	57	<p style="text-align: center;"> Dr. Stephen Brooks steve.brooks@hc-sc.gc.ca (603) 957-0932 </p> <p style="text-align: center;"> Kurtis Sarafin kurtis.sarafin@hc-sc.gc.ca (613) 957-0921 </p>
				16.7 - 162	Q1 2023	70	
				16.7 - 157	Q4 2022	68	
				16.7 - 149	Q3 2022	75	
				17.4 - 211	A2 2022	72	
				17.4 - 211	Q1 2022	68	
				17.4 - 211	Q4 2021	72	
				18.1 - 211	Q3 2021	78	
				13.1 - 183	Q2 2021	80	
				13.1 - 220	Q1 2021	72	
				13.1 - 220	Q4 2020	62	
				13.1 - 220	Q3 2020	50	
				16.1 - 220	Q2 2020	40	
				16.7 - 140	Q1 2020	45	
19.7 - 140	Q4 2019	45					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Nutritional Biomarkers Branch CDC Atlanta, GA	LC-MS/MS ⁺	#4027 routine method for serum 25OHD3, 25OHD2, C3-epimer-25OHD3	4 - 430	12.1 - 208	Q3 2023	82	Rosemary Schleicher, PhD RSchleicher@cdc.gov (770) 488-4424
				12.1 - 158	Q2 2023	85	
				12.1 - 139	Q1 2023	90	
				16.7 - 149	Q4 2022	90	
				16.7 - 158	Q3 2022	90	
				16.7 - 158	Q2 2022	90	
				17.4 - 158	Q1 2022	92	
				22.0 - 158	Q4 2021	82	
				17.4 - 157	Q3 2021	78	
				17.4 - 140	Q2 2021	80	
				13.1 - 140	Q1 2021	72	
				13.1 - 140	Q4 2020	80	
				13.1 - 140	Q3 2020	78	
				13.1 - 140	Q2 2020	70	
19.7 - 140	Q1 2020	65					
19.7 - 183	Q4 2019	70					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Ortho Clinical Diagnostics Pencoed, Wales, UK	CIA	Vitros	21.6 - 315	36.4 - 182	Q3 2023	15	Aled Yates aled.yates@orthoclinicaldiagnostics.com (+44) 1656 778397
				36.4 - 162	Q1 2023	23	
				36.4 - 162	Q4 2022	23	
				36.4 - 197	Q3 2022	18	
				36.7 - 211	Q2 2022	13	
				36.7 - 211	Q1 2022	10	
				36.4 - 211	Q4 2021	21	
				36.4 - 211	Q3 2021	21	
				36.4 - 123	Q2 2021	20	
				27.5 - 134	Q1 2021	15	
				22.6 - 136	Q4 2020	15	
				10.4 - 135	Q2 2020	15	
				10.4 - 140	Q1 2020	18	
19.7 - 220	Q4 2019	22					
Penn State University, Hershey Medical Center, Dept. of Pathology and Laboratory Medicine Hershey, PA	LC/MS/MS†	25OHD2 and 25OHD3 (39D0657304)	9.98 - 399	16.7 - 147	Q4 2019	62	Yusheng Zhu, PhD, DABCC, FAACC yzhu@pennstatehealth.psu.edu (717) 731-5123
PerkinElmer Health Sciences, Inc Hayward, CA	ELISA	25 OH Vitamin D ELISA 80min	4.18 - 375	17.4 - 198	Q4 2019	28	Shawn You Shaofang.you@perkinelmer.com (650) 872-1800

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Quest Diagnostics Nichols Institute of Valencia, Inc Valencia, California	LC/MS/MS+	Total 25-Hydroxyvitamin D (Vitamin D2/Vitamin D3)	10 - 1,280	17.4 - 162	Q1 2023	57	Amit Ghoshal PhD Amit.K.Ghoshal@QuestDiagnostics.com Quest Diagnostics Nichols Institute Valencia, Inc. (661) 799-6204
				16.1 - 182	Q1 2022	25	
				12.1 - 158	Q4 2021	42	
				12.1 - 158	Q3 2021	35	
				12.1 - 123	Q2 2021	42	
				12.1 - 220	Q1 2021	50	
				16.7 - 220	Q3 2020	55	
				16.7 - 220	Q2 2020	65	
				18.1 - 149	Q1 2020	62	
				18.1 - 140	Q4 2019	62	
Roche Diagnostics GmbH Penzberg, DE	ECL binding assay	Elecsys Vitamin D total II	7.38 - 885	16.1 - 182	Q3 2023	92	Christian Vogl christian.vogl.cv1@roche.com +49 8856 60 19758
				16.7 - 208	Q3 2022	28	
				16.7 - 182	Q2 2022	20	
				16.1 - 183	Q1 2022	20	
				16.1 - 183	Q4 2021	20	
				22.0 - 183	Q1 2021	42	
				16.7 - 134	Q4 2020	40	
				13.1 - 136	Q3 2020	42	
				13.1 - 198	Q2 2020	38	
				13.1 - 198	Q1 2020	45	
13.1 - 198	Q4 2019	45					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Roche Diagnostics GmbH Penzberg, DE	ECL binding assay	Elecsys Vitamin D total III	7.5 - 300	12.1 - 158 17.4 - 158 18.1 - 158 18.1 - 162 17.4 - 208 17.4 - 208 17.4 - 208 12.1 - 208 12.1 - 149 12.1 - 149	Q2 2023 Q1 2023 Q4 2022 Q3 2022 Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q4 2020	42 40 40 38 28 35 32 22 15 25	Christian Vogl christian.vogl.cv1@roche.com +49 8856 60 19758
Roche Diagnostics GmbH Penzberg, DE	LC/MS/MS+	25-Hydroxyvitamin D3 and 25-Hydroxyvitamin D2 in serum	D3: 3.75 - 450 D2: 3.64 - 437	12.1 - 182 12.1 - 182 12.1 - 182 12.1 - 162 17.4 - 182 18.1 - 182 21.5 - 182 16.1 - 182 16.1 - 134 16.1 - 134	Q2 2023 Q1 2023 Q4 2022 Q3 2022 Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021	92 90 90 80 70 68 68 72 78 65	Judith Taibon judith.taibon@roche.com 0049 8856 60 12941
SCIEX Diagnostics Framingham, MA	LC/MS/MS+	Total 25-Hydroxyvitamin D (Vitamin D2/Vitamin D3)	10 - 800	12.1 - 183 19.7 - 183	Q1 2020 Q4 2019	55 52	Theresa Irvine Theresa.Irvine@sciex.com (508) 383-7409

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Siemens Healthcare Diagnostics Tarrytown, NY	Chemiluminescence Immunoassay	ADVIA Centaur® Vitamin D Total assay	10.5 - 375	16.1 - 158	Q3 2023	32	Neil Parker neil.np.parker@siemens-healthineers.com
				12.1 - 162	Q2 2023	28	
				12.1 - 162	Q1 2023	28	
				12.1 - 208	Q4 2022	25	
				16.1 - 208	Q2 2022	25	
				16.1 - 208	Q1 2022	23	
				17.4 - 197	Q4 2021	26	
				17.4 - 197	Q3 2021	15	
				13.1 - 135	Q2 2021	18	
				13.1 - 149	Q1 2021	18	
				13.1 - 149	Q4 2020	18	
				13.1 - 149	Q3 2020	20	
				16.1 - 149	Q2 2020	15	
				16.1 - 147	Q1 2020	15	
17.4 - 183	Q4 2019	22					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Siemens Healthcare Diagnostics Tarrytown, NY	Chemiluminescence Immunoassay	Atellica® IM Vitamin D Total assay	10.5 - 375	22.2 - 162	Q3 2023	28	Neil Parker neil.np.parker@siemens-healthineers.com
				12.1 - 158	Q2 2023	38	
				12.1 - 158	Q1 2023	35	
				12.1 - 208	Q4 2022	30	
				12.1 - 208	Q3 2022	30	
				17.4 - 208	Q2 2022	32	
				12.1 - 208	Q1 2022	38	
				12.1 - 162	Q4 2021	38	
				12.1 - 162	Q3 2021	36	
				12.1 - 132	Q2 2021	18	
				12.1 - 132	Q1 2021	28	
				12.1 - 132	Q4 2020	31	
				12.1 - 134	Q3 2020	32	
				12.1 - 220	Q2 2020	38	
				16.7 - 220	Q1 2020	35	
19.7 - 220	Q4 2019	30					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Siemens Healthcare Diagnostics Glasgow, DE	Chemiluminescence Immunoassay	Dimension®ExL™ Vitamin D Total assay	12.5 - 375	22.2 - 208	Q3 2023	35	Dr. Tie Wei tie.g.wei@siemens-healthineers.com (302) 631-6929 Jie Li jie.l.li@siemens-healthineers.com (302) 631-7708
				22.0 - 208	Q2 2023	44	
				16.7 - 208	Q1 2023	44	
				16.7 - 208	Q4 2022	44	
				16.7 - 157	Q3 2022	41	
				16.7 - 162	Q2 2022	40	
				22.2 - 162	Q1 2022	50	
				18.1 - 162	Q4 2021	45	
				18.1 - 162	Q3 2021	48	
				18.1 - 140	Q2 2021	50	
				16.7 - 140	Q4 2020	38	
				12.1 - 140	Q3 2020	38	
				12.1 - 175	Q2 2020	35	
				10.4 - 183	Q1 2020	38	
10.4 - 183	Q4 2019	32					
SNIBE Shenzhen, China	LC/MS/MS†	25-OH Vitamin D	12 - 250	12.1-149	Q2 2023	65	Huijun Cui huijun.cui@snibe.cn 0086 18171423440
SNIBE Shenzhen, China	Chemiluminescence Immunoassay	Maglumi 25-OH Vitamin D	7.5 - 375	12.1 - 149 22.0 - 230	Q4 2022 Q4 2019	32 28	Jinyun Yuan jinyun.yuan@snibe.cn Huijun Cui huijun.cui@snibe.cn

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
Thermo Fisher Scientific Vantaa, Finland	LC/MS/MS ⁺	Thermo Scientific™ Cascadion™ SM 25-Hydroxy Vitamin D	8.24 - 649.3	16.1 - 149	Q3 2022	80	Sirpa Riistama-Laari Sirpa.riistama-laari@thermofisher.com +358 103292215
				16.1 - 157	Q2 2022	85	
				16.1 - 157	Q1 2022	82	
				18.1 - 157	Q4 2021	82	
				18.1 - 157	Q3 2021	82	
				18.1 - 149	Q2 2021	80	
				18.1 - 149	Q1 2021	80	
				13.1 - 135	Q4 2020	75	
				13.1 - 147	Q3 2020	78	
				13.1 - 147	Q2 2020	75	
				13.1 - 220	Q1 2020	75	
16.1 - 220	Q4 2019	80					
Tosoh Corporation Tokyo, Japan	Enzyme Immunoassay	ST AIA-PACK 25- OH Vitamin D	10 - 375	12.1 - 208	Q2 2023	18	Dr. Sachi-yuki Hasegawa sachi-yuki-hasegawa-de@tosoh.co.jp Hisao Tsukamoto hisao-tsukamoto-rn@tosoh.co.jp +81-467-76-5384
				16.1 - 162	Q1 2023	13	
				16.1 - 149	Q4 2022	15	
				16.1 - 182	Q3 2022	12	
				12.1 - 181	Q3 2021	15	
				12.1 - 128	Q2 2021	22	
				12.1 - 149	Q1 2021	20	
				16.7 - 149	Q4 2020	20	
				16.7 - 183	Q3 2020	18	
				16.7 - 198	Q2 2020	12	
				19.7 - 198	Q1 2020	20	
19.7 - 198	Q4 2019	28					

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (nmol/L)	Certification Measurement Range (nmol/L)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
University Hospital – Biochemistry Laboratory Grenoble, France	LC/MS/MS†	LC/MS/MS† assay 2D LC METHODE	4 - 360	13.1 - 183	Q1 2020	72	Anne Sophie Gauchez ASGauchez@chu-grenoble.fr SGazzah@chu-grenoble.fr
University of Minnesota, Molecular Epidemiology and Biomarker Research Laboratory (MEBRL) Minneapolis, MN	LC/MS/MS†	25-Hydroxy Vitamin D2 and D3	5.3 - 1,240	12.1 - 197 16.1 - 197 16.1 - 197 16.1 - 149 13.1 - 136 12.1 - 140 12.1 - 183	Q2 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021 Q2 2020 Q4 2019	40 52 52 60 55 42 45	Dr. Myron Gross gross001@umn.edu (612) 624-5417
Waters Technologies Wexford, Ireland	LC/MS/MS†	Total 25-Hydroxyvitamin D (Vitamin D2/Vitamin D3)	20 - 750	12.1 - 208 22.0 - 162 27.5 - 211 22.2 - 211 22.2 - 211 22.0 - 211 22.0 - 140 13.1 - 140 13.1 - 140 13.1 - 140	Q3 2023 Q3 2022 Q2 2022 Q1 2022 Q4 2021 Q3 2021 Q2 2021 Q1 2021 Q4 2020 Q3 2020	85 50 57 60 57 57 65 60 60 55	Norma Breen Norma_Breen@waters.com 00353 (0)53-9160514

† LC/MS/MS – Liquid Chromatography Tandem Mass Spectrometry

* Certification is only provided for serum. No assessment was performed in plasma.