

anti-HIV Antibodies (HIV)

Set 2 - June 2021

Analysis Method: Avioq HIV-1 Microelisa

Expiration Date: 7/31/2022

Material Storage: -20°C ± 10°C

Units: OD

Lot	Level	Analyte	Mean Value	95% LL	95% UL
A2007	Negative	anti-HIV Antibodies (HIV)	0.135	0.072	0.198
B2007	Low Positive	anti-HIV Antibodies (HIV)	1.578	1.035	2.121
C2007	Low Positive	anti-HIV Antibodies (HIV)	1.076	0.618	1.533
D2007	High Positive	anti-HIV Antibodies (HIV)	2.174	1.616	2.731

Analysis Method: Western Blot, Band ID and Serostatus

Lot	Level	gp160	gp120	p66	p55/51	Gp41	P40	P31	P24	P18
A2007	Negative	-	-	-	-	+/-	-	-	-	-
B2007	Low Positive	++	+/-	+	+	+/-	++	++	++	+/-
C2007	Low Positive	++	-	+/-	+	+/-	++	+	++	-
D2007	High Positive	++	+	++	++	++	++	++	++	++