

Immunoreactive Trypsinogen (IRT)

Set 2 - June 2021

Analysis Method: Perkin Elmer GSP Neonatal Kit

Expiration Date: Sep. 2022

Material Storage: -20°C ± 10°C

Units: ng/mL blood

| Lot | Analyte | Assayed Value | 95% LL | 95% UL |
|-------|----------------------------------|---------------|--------|--------|
| A2009 | Immunoreactive Trypsinogen (IRT) | 19.4 | 17.3 | 21.5 |
| B2009 | Immunoreactive Trypsinogen (IRT) | 63.9 | 56.5 | 71.3 |
| C2009 | Immunoreactive Trypsinogen (IRT) | 129.3 | 116.4 | 142.1 |
| D2009 | Immunoreactive Trypsinogen (IRT) | 231.3 | 205.8 | 256.8 |