

Thyroxine (T4)

Set 2 - June 2021

Analysis Method: Perkin Elmer GSP Neonatal

Expiration Date: December 2022

Material Storage: -20°C ± 10°C

Units: µg/dL serum

Lot	Analyte	Enrichment	Mean Value	95% LL	95% UL
A2000	Thyroxine (T4)	2	1.5	0.9	2.0
B2000	Thyroxine (T4)	7	7.1	5.9	8.3
C2000	Thyroxine (T4)	11	11.4	9.5	13.2