

Newborn Screening Quality Assurance Program
 Lysophosphatidylcholines (XALD) Quality Control Specimen Certification
 Set 1—January 2016

EXPIRATION: May 2017

Analysis Method: LC-MS/MS

ENRICHMENT LEVELS (ENDOGENOUS LEVELS NOT INCLUDED)

<i>Analyte ($\mu\text{mol/L}$ whole blood)</i>	<i>Lot</i>	<i>Base/Low</i>	<i>Lot</i>	<i>Low</i>	<i>Lot</i>	<i>High</i>
24-LPC*	A1510	0.0	B1510	1.0	C1510	5.0
26-LPC**	A1510	0.0	B1510	1.0	C1510	5.0

ANALYTICAL INFORMATION

<i>Lot Numbers, Mean Values (\bar{x}), and 95% Confidence Limits (CL)</i>						
<i>Analyte</i>	<i>Lot</i>	<i>Base/Low</i>	<i>Lot</i>	<i>Low</i>	<i>Lot</i>	<i>High</i>
24-LPC*	A1510	$\bar{x} = 0.04$ CL = 0.02–0.05	B1510	$\bar{x} = 0.44$ CL = 0.29–0.59	C1510	$\bar{x} = 2.18$ CL = 1.46–2.90
26-LPC**	A1510	$\bar{x} = 0.03$ CL = 0.02–0.05	B1510	$\bar{x} = 0.86$ CL = 0.54–1.19	C1510	$\bar{x} = 4.58$ CL = 3.34–5.81

*Tetracosanoyl lysophosphatidylcholine

**Hexacosanoyl lysophosphatidylcholine

Note: The values provided in the above tables are for reference use only. The mean value and confidence limits (CL) are determined by CDC for each Quality Control (QC) lot. Each participating laboratory must establish its own mean values and CL for its test method with these QC materials. Temporary estimates of mean values and CL can be determined after 10 successive, independent measurements. *Slazyk WE, Hannon WH. Quality Assurance in the Newborn Screening Laboratory. In: Therrell BL Jr, editor. Laboratory methods for neonatal screening. Washington (DC): American Public Health Association, 1993:23-46.*