

Title: Constant vs. cyclic flow when testing face masks and respirators as source control devices for simulated respiratory aerosols—Data Dictionary

Dataset Number:

Data dictionary

| <i>Field name on data page</i> | <i>Field definition</i> |
|--------------------------------|--|
| Source_control_device | Type of source control device used in experiment |
| Flow_rate | Airflow (cyclic or constant) used in experiment in Liters/minute |
| Flow_type | Type of airflow (cyclic flow or constant flow) |
| Replicate | Each combination of parameters was tested four times |
| Fit_factor | Reading from PortaCount respirator fit tester |
| XXX_μm_particles_per_cm^3 | Average concentration of aerosol particles within each size bin in number of particles/cm ³ . For example, 0.337_μm_particles_per_cm^3 is the number of aerosol particles with a diameter of 0.337 micrometers per cubic centimeter in the test chamber. The particle diameter refers to the arithmetic mean of the optical diameter for each size bin. |
| Filtration_efficiency | Filtration efficiency results for each source control device in units of fraction of the test aerosol (dimensionless). |
| Filter_resistance | Airflow resistance results for each source control device in units of Pascals. |