

**Miller, Diane M.**

---

**From:** jcalpin@analyticscorp.com  
**Sent:** Thursday, December 22, 2005 2:54 PM  
**To:** sa-cin-titanium  
**Subject:** Titanium Dioxide Comments

**Follow Up Flag:** Follow up  
**Flag Status:** Completed

**Name**  
James Calpin

**Organization**  
Analytics Corp

**Email**  
jcalpin@analyticscorp.com

**Address**  
8040 Villa Park Drive  
Richmond, VA 23228  
United States

**Comments**  
Referring to lines 2043 & 2044, as an AIHA accredited laboratory director, I do not agree that NIOSH 7300 be the refenced method for TiO<sub>2</sub> since it has poor solubility in the acids used in the filter digestion. OSHA ID 125 is a better choice for the referenced analytical method.  
James Calpin, MS, C.I.H.