



## LEAD by Flame AAS

7082

Pb MW: 207.19 (Pb) CAS: 7439-92-1 (Pb) RTECS: OF7525000 (Pb)  
 MW: 223.19 (PbO) CAS: 1317-36-8 (PbO) RTECS: OG1750000 (PbO)

**METHOD:** 7082, Issue 3

**EVALUATION:** FULL

**Issue 1:** 15 February 1984

**Issue 3:** 12 July 2017

**OSHA:** 0.050mg/m<sup>3</sup>  
**NIOSH:** 0.050 mg/m<sup>3</sup>  
**OTHER OELs:** [1-3]

**PROPERTIES:** soft metal; d 11.3 g/cm<sup>3</sup>; MP 327.5 °C  
 valences +2, +4 in salts

**SYNONYMS:** elemental lead and lead compounds, except alkyl lead

SAMPLING		MEASUREMENT	
<b>SAMPLER:</b>	FILTER (0.8-µm cellulose ester membrane) or INTERNAL CAPSULE, cellulose acetate dome with inlet opening attached to filter	<b>TECHNIQUE:</b>	ATOMIC ABSORPTION SPECTROPHOTOMETER, FLAME
<b>FLOW RATE:</b>	1 to 4 L/min	<b>ANALYTE:</b>	lead
<b>VOL-MIN:</b>	200 L @ 0.05 mg/m <sup>3</sup>	<b>ASHING:</b>	conc. HNO <sub>3</sub> , 6 mL + 30% H <sub>2</sub> O <sub>2</sub> , 1 mL; 140°C
<b>-MAX:</b>	1500 L	<b>FINAL SOLUTION:</b>	10% HNO <sub>3</sub> , 10 mL
<b>SHIPMENT:</b>	routine	<b>FLAME:</b>	air-acetylene, oxidizing
<b>SAMPLE STABILITY:</b>	stable at least 7 weeks [4]	<b>WAVELENGTH:</b>	283.3 nm
<b>BLANKS:</b>	2 to 10 field blanks per set	<b>BACKGROUND CORRECTION:</b>	D <sub>2</sub> or H <sub>2</sub> lamp, or Zeeman
ACCURACY		<b>CALIBRATION:</b>	Pb <sup>2+</sup> in 10% HNO <sub>3</sub>
<b>RANGE STUDIED:</b>	0.13 to 1.7 mg/m <sup>3</sup> [8]	<b>RANGE:</b>	10 to 200 µg/sample [6,7]
<b>BIAS:</b>	-3.1%	<b>ESTIMATED LOD:</b>	2.6 µg/sample [8]
<b>OVERALL PRECISION (<math>\hat{S}_{r,T}</math>):</b>	0.07 [5,6]	<b>PRECISION (<math>\bar{S}_r</math>):</b>	0.03 [5]
<b>ACCURACY:</b>	±17.6%		

**APPLICABILITY:** The working range is 0.05 to >1 mg/m<sup>3</sup> for a 200-L air sample. The method is applicable to elemental lead, including Pb fume, and all other aerosols containing lead. This is an elemental analysis, not compound specific. Aliquots of the samples can be analyzed separately for additional elements. This method has been updated to include internal capsule samplers.

**INTERFERENCES:** Use D<sub>2</sub> or H<sub>2</sub> continuum or Zeeman background correction to control flame or molecular absorption. High concentrations of calcium, sulfate, carbonate, phosphate, iodide, fluoride, or acetate can be corrected for.

**OTHER METHODS:** This method combines and replaces P&CAM 173 [7] and S341 [8] for lead. NIOSH Methods 7300 (ICP-AES), 7701 (ASV) and 7105 (GFAAS) are alternative analytical methods. NIOSH Method 7505 is specific for lead sulfide. A consensus standard method, ASTM D6785, has been published [9].

**REAGENTS:**

1. Nitric acid, conc., trace metal grade\*
2. Nitric acid, 10% (v/v). Add 100 mL conc. HNO<sub>3</sub> to 500 mL water; dilute to 1 L
3. Hydrogen peroxide, 30% H<sub>2</sub>O<sub>2</sub> (w/w), reagent grade.\*
4. Calibration stock solution, 1000 µg/mL Pb. Commercial standard or dissolve 1.00 g Pb metal in minimum volume of (1+1) HCl and dilute to 1 L with 1% (v/v) HCl. Store in a polyethylene bottle. Stable > one year.
5. Air, compressed, filtered.
6. Acetylene
7. Distilled or deionized water

\*See SPECIAL PRECAUTIONS.

**EQUIPMENT:**

1. Sampler: Cellulose ester filter or cellulose acetate internal capsule attached to mixed cellulose ester membrane filter, 0.8-µm pore size, 37-mm diameter, in 2-piece cassette filter holder.
2. Personal sampling pump, 1 to 4 L/min, with flexible connecting tubing.
3. Atomic absorption spectrophotometer with an air-acetylene burner head and background correction.
4. Lead hollow cathode lamp or electrode dischargeless lamp.
5. Regulators, two-stage, for air and acetylene.
6. Beakers, Phillips, 125-mL, or Griffin, 50-mL with watch glass covers.\*\*
7. Volumetric flasks, 10- and 100-mL.\*\*
8. Assorted volumetric pipets, as needed.\*\*
9. Hotplate, surface temperature 140 °C.
10. Bottles, polyethylene, 100-mL.

\*\* Clean all glassware with conc. nitric acid and rinse thoroughly with distilled or deionized water before use.

---

**SPECIAL PRECAUTIONS:** Wear appropriate personal protection during sampling activities and analysis. It is essential that suitable gloves, eye protection, laboratory coat, etc., be used when working with the chemicals. Concentrated nitric acid is an irritant and may burn skin. Perform all acid digestions in a fume hood. Hydrogen peroxide is a strong oxidizing agent, a strong irritant, and corrosive to the skin. Wear gloves and eye protection.

---

**SAMPLING:**

1. Calibrate each personal sampling pump with a representative sampler in line.  
NOTE: See NMAM guidance chapters for discussion on sampling.
2. Sample at an accurately known flow rate between 1 and 4 L/min (± 5%) for up to 8 h for a total sample size of 200 to 1500 L for TWA measurements. Do not exceed a filter loading of approximately 5 mg total dust.  
NOTE: Filter overloading can be assessed by periodic visual checks. See NMAM guidance chapters for additional discussion on filter capacity.

**SAMPLE PREPARATION:**

3. Open the cassette filter holders and transfer the samples and blanks to clean beakers.  
NOTE: If internal capsules are not used, wipe the internal cassette surfaces with a polyvinyl alcohol wipe or cellulosic wipe wetted with deionized water, and add the wipe to the digestion vessel (to transfer non-filter aerosol deposits into the digestion vessels) [10].

NOTE: The following sample preparation gave quantitative recovery (see EVALUATION OF METHOD) [8]. Steps 4 through 9 of Method 7300 or other quantitative ashing techniques may be substituted, especially if several metals are to be determined on a single filter.

NOTE: The Appendix gives a microwave digestion procedure, which may be necessary for complete recovery of lead from some matrices, especially epoxy-based paint.

4. Add 3 mL conc. HNO<sub>3</sub>, and 1 mL 30% H<sub>2</sub>O<sub>2</sub> and cover with a watch glass. Start reagent blanks at this step.

NOTE: If PbO<sub>2</sub> is not present in the sample, the 30% H<sub>2</sub>O<sub>2</sub> need not be added [6,8].

5. Heat on 140 °C hotplate until volume is reduced to about 0.5 mL.
6. Repeat two more times using 2 mL conc. HNO<sub>3</sub> and 1 mL 30% H<sub>2</sub>O<sub>2</sub> each time.
7. Heat on 140 °C hotplate until ca. 0.5 mL liquid remains.
8. When sample is dry, rinse the watch glass and walls of the beaker with 3 to 5 mL 10% HNO<sub>3</sub>. Allow the solution to evaporate to dryness.
9. Cool each beaker and dissolve the residues in 1 mL conc. HNO<sub>3</sub>.
10. Transfer the solution quantitatively to a 10-mL volumetric flask and dilute to volume with distilled water.

NOTE: If the concentration (M) of any of the following is expected to exceed the lead concentration (M) by 10-fold or more, add 1 mL 1 M Na<sub>2</sub>EDTA to each flask before dilution to volume: CO<sub>3</sub><sup>2-</sup>, PO<sub>4</sub><sup>3-</sup>, I<sup>-</sup>, F<sup>-</sup>, CH<sub>3</sub>COO<sup>-</sup> [8]. If Ca<sup>2+</sup> or SO<sub>4</sub><sup>2-</sup> are present in 10-fold or greater excess, make all standards and samples with 1% (w/w) La<sup>2+</sup> [7].

#### CALIBRATION AND QUALITY CONTROL:

11. Prepare a series of working standards covering the range 0.25 to 20 µg/mL Pb (2.5 to 200 µg Pb per sample).
  - a. Add aliquots of calibration stock solution to 100-mL volumetric flasks. Dilute to volume with 10% HNO<sub>3</sub>. Store the working standards in polyethylene bottles and prepare fresh weekly.
  - b. Analyze the working standards together with the blanks and samples (steps 14 and 15).
  - c. Prepare a calibration graph of absorbance vs. solution concentration (µg/mL)
12. Aspirate a standard for every 10 samples to check for instrument drift.
13. Check recoveries with at least one spiked media blank per 10 samples. Use method of standard additions occasionally to check for interferences.

#### MEASUREMENT:

14. Set spectrophotometer as specified by the manufacturer and to conditions on page 7082-1.
 

NOTE: An alternate wavelength is 217.0 nm [11]. Analyses at 217.0 nm have slightly greater sensitivity, but poorer signal-to-noise ratio compared to 283.3 nm. Also, non-atomic absorption is significantly greater at 217.0 nm, making the use of D<sub>2</sub> or H<sub>2</sub> continuum, or Zeeman background correction mandatory at that wavelength.
15. Aspirate standards, samples, and blanks. Record absorbance readings.
 

NOTE: If the absorbance values for the samples are above the linear range of the standards, dilute with 10% HNO<sub>3</sub>, reanalyze, and apply the appropriate dilution factor in the calculations.

#### CALCULATIONS:

16. Using the measured absorbances, calculate the corresponding concentrations (µg/mL) of lead in the sample, C<sub>s</sub>, and average media blank, C<sub>b</sub>, from the calibration graph.
17. Using the solution volumes (mL) of the sample, V<sub>s</sub>, and media blanks, V<sub>b</sub>, calculate the concentration, C (mg/m<sup>3</sup>), of lead in the air volume sampled, V (L):

$$C = \frac{C_s V_s - C_b V_b}{V}, \frac{mg}{m^3}$$

NOTE: µg/mL ≈ mg/m<sup>3</sup>.

**EVALUATION OF METHOD:**

The predecessor to NIOSH 7082, Method S341 [8], was issued on October 24, 1975, and validated over the range 0.13 to 0.4 mg/m<sup>3</sup> for a 180-L air sample, using generated atmospheres of lead nitrate [5]. Recovery in the range 18 to 72 µg Pb per sample was 98%, and collection efficiency of 0.8-µm mixed cellulose ester filters (Millipore Type AA) was 100% for the aerosols at the detection limit of 0.013 mg/m<sup>3</sup>. Subsequent studies on analytical recovery of 200 µg Pb per sample gave the following results [6,8]:

Species	Digestion method	Analytical recovery, %
Pb metal	HNO <sub>3</sub> only	92.4 ± 4
Pb metal	HNO <sub>3</sub> + H <sub>2</sub> O <sub>2</sub>	103 ± 3
PbO	HNO <sub>3</sub> only	93 ± 4
PbS	HNO <sub>3</sub> only	93 ± 5
PbO <sub>2</sub>	HNO <sub>3</sub> only	82 ± 3
PbO <sub>2</sub>	HNO <sub>3</sub> + H <sub>2</sub> O <sub>2</sub>	100 ± 1
Pb in paint*	HNO <sub>3</sub> only	95 ± 6
Pb in paint*	HNO <sub>3</sub> + H <sub>2</sub> O <sub>2</sub>	95 ± 6

\*Standard Reference Material #1579, U.S. National Institute of Standards and Technology

Additional collection efficiency studies were also done using Gelman GN-4 metrical cellulose acetate membrane filters for the collection of Pb fume, which had geometric mean diameter of 0.1 µm [5]. Mean collection efficiency for 24 sampling runs at flow rates between 0.15 and 4.0 L/min was >97 ± 2%. Overall precision,  $S_{rT}$ , was 0.072 for lead nitrate aerosol [5,8] and 0.068 for Pb fume [6,8].

Evaluation information on internal capsule samplers may be found in NIOSH 7306 [12].

**REFERENCES:**

- [1] ACGIH [2016]. TLVs<sup>®</sup> and BEIs<sup>®</sup> based on the documentation of the Threshold Limit Values for chemical substances and physical agents and Biological Exposure Indices. Cincinnati, Ohio: American Conference of Governmental Industrial Hygienists.
- [2] Permanent Senate Commission for the Investigation of Health Hazards of Chemical Compounds in the Work Area [2016]. List of MAK and BAT values 2016; Maximum Concentrations and Biological Tolerance Values at the Workplace. Weinheim, Germany: Wiley-VCH Verlag GmbH & Co., <http://dx.doi.org/10.1002/9783527805983>.
- [3] Institute for Occupational Safety and Health of the German Social Accident Insurance (IFA) [no date]. GESTIS international limit values, <http://limitvalue.ifa.dguv.de/>. Date accessed: July 12, 2017.
- [4] Marks GE, Knutsen EO [1975]. Complete testing of the NIOSH method for the determination of trace metals by atomic absorption spectrophotometry. Chicago, IL: IIT Research Institute, NIOSH contract no. CDC-99-74-48.
- [5] Taylor DG, Kupel RE, Bryant JM. [1977]. Documentation of the NIOSH validation tests. Cincinnati, OH: U.S. Department of Health, Education, and Welfare, Center for Disease Control, National Institute for Occupational Safety and Health, DHEW (NIOSH) Publication No. 77-185. Available from [www.ntis.gov](http://www.ntis.gov) as accession no. PB274248
- [6] Gutknecht WF, Ranade MH, Grohse PM, Damle A, O'Neal D [1981]. Heavy metal aerosols: Collection and dissolution efficiencies. Research Triangle Park, NC: Research Triangle Institute, Contract 210-79-0058, Available from [www.ntis.gov](http://www.ntis.gov) as accession no. PB83-106740.

- 
- [7] NIOSH [1979]. Method P&CAM 173. In: Taylor DG, ed. NIOSH manual of analytical methods. 2<sup>nd</sup> ed. Cincinnati, OH: U.S. Department of Health, Education and Welfare, Center for Disease Control, National Institute for Occupational Safety and Health, DHEW (NIOSH) Publication No. 77-157-A.
- [8] NIOSH [1982]. Method S341. In: Taylor DG, ed. NIOSH manual of analytical methods. 2<sup>nd</sup> ed. Cincinnati, OH: U.S. Department of Health and Human Services, Center for Disease Control, National Institute for Occupational Safety and Health, DHEW (NIOSH) Publication No. 82-100.
- [9] ASTM International [2013]. ASTM D6785, Standard test method for determination of lead in workplace air using flame or graphite furnace atomic absorption spectrometry. West Conshohocken, PA: ASTM International.
- [10] Ashley K, Harper M [2013]. Closed-face filter cassette (CFC) sampling - Guidance of procedures for inclusion of material adhering to internal sampler surfaces. *J Occup Environ Hyg* 10:D29-D33.
- [11] Perkin-Elmer Corporation [1976]. Analytical methods for atomic absorption spectrophotometry. Norwalk, CT: Perkin-Elmer Corporation.
- [12] NIOSH [2015]. Method 7306. In: Ashley K, O'Connor PF, eds. NIOSH manual of analytical methods. 5<sup>th</sup> ed. Cincinnati, OH: U.S. Department of Health and Human Services, Centers for Disease Control and Prevention, National Institute for Occupational Safety and Health, DHHS (NIOSH) Publication No. 14-151.
- [13] DataChem Laboratories [no date]. In-house procedure for microwave sample digestion. Salt Lake City, UT: DataChem Laboratories. Unpublished.
- [14] EPA [1986]. SW-846: Test methods for evaluating solid waste, Physical and chemical methods, 3<sup>rd</sup> ed. Washington, DC: U.S. Environmental Protection Agency.
- [15] Kingston HM, Jassie LB, eds. [1988]. Safety guidelines for microwave systems in the analytical laboratory. In: Introduction to microwave acid decomposition: Theory and practice. Washington, DC: American Chemical Society.
- [16] ASTM International [2011]. ASTM D1193, Standard specification for reagent water. West Conshohocken, PA: ASTM International.
- [17] Kingston HM, Jassie LB, eds. [1988]. Introduction to microwave acid decomposition: Theory and practice. Washington, DC: American Chemical Society.
- [18] Kingston HM [1988]. Quarterly Report, EPA IAG #DW1-393254-01-1.
- [19] Binstock DA, Yeager WM, Grohse PM, Gaskill A [1989]. Validation of a method for determining elements in solid waste by microwave digestion, Draft technical report for RTI Project number 321U-3579-24. Research Triangle Park, NC: Research Triangle Institute.

**METHOD REVISED BY:**

Mark Millson, R. DeLon Hull, PhD, and Ronnee N. Andrews, PhD, NIOSH. S341 originally validated under NIOSH contract CDC-94-74-45; additional studies under NIOSH contract 210-79-0058.

James B. Perkins, David L. Wheeler, and Keith Nicholson, Ph.D., DataChem Laboratories, Salt Lake City, UT, prepared the microwave digestion procedure described in the Appendix.

**Disclaimer:** Mention of any company or product does not constitute endorsement by NIOSH. In addition, citations to websites external to NIOSH do not constitute NIOSH endorsement of the sponsoring organizations or their programs or products. Furthermore, NIOSH is not responsible for the content of these websites. All web addresses referenced in this document were accessible as of the publication date.

## APPENDIX: MICROWAVE DIGESTION FOR LEAD IN PAINT CHIPS (AND OTHER MATRICES)

This procedure is an alternative to the procedure presented in the Sample Preparation section of this method. It provides a rapid, complete acid digestion prior to analysis by flame atomic absorption (FAA), heated graphite furnace atomic absorption (HGFAA), and inductively coupled plasma spectroscopy (ICP) [13].

### Apparatus and Materials [14-19]

1. Microwave unit, to provide programmable power with a minimum of 574 W and programmable to within  $\pm 10$  W of the required power.
2. The microwave unit cavity shall be corrosion resistant as well as ventilated. All electronics are protected against corrosion for safe operation.
3. The system requires microwave transparent and reagent resistant vessels, such as perfluoroalkoxy alkane (PFA) digestion vessels (120-mL capacity), capable of withstanding pressures up to  $7.5 \pm 0.7$  atm ( $760 \pm 70$  kPa). Vessels shall also be capable of controlled pressure relief at pressures exceeding  $7.5 \pm 0.7$  atm ( $760 \pm 70$  kPa). Other, equivalent types of vessels designed to operate at temperatures and pressures required and recommended by manufacturer can be used.
4. A rotating turntable is employed to ensure homogeneous distribution of microwave radiation within the unit. The speed of the turntable should be a minimum of 3 rpm.
5. A safety concern relates to the use of sealed containers without pressure relief valves in the unit. Temperature is the important variable controlling the reaction. Pressure is needed to attain elevated temperatures but must be safely contained [15].
6. Polymeric volumetric ware in plastic (PTFE or polyethylene), 50- or 100-mL capacity.
7. Disposable polypropylene filter funnel.
8. Analytical balance, 300-g capacity, and minimum  $\pm 0.001$  g.

### Reagents

1. Nitric acid, concentrated, spectroscopy grade.
2. Reagent Water. Reagent water shall be interference free. All references to water in the method refer to reagent water that meets the ASTM Type 2 standard.

### Procedure

1. Calibration of Microwave Equipment
  - a. Calibrate microwave equipment in accordance with manufacturer's instructions. If calibration instructions are not available, see EPA Method 3051 [14].
2. All digestion vessels and volumetric ware must be carefully acid washed and rinsed with reagent water. All digestion vessels should be cleaned by leaching with hot (1:1) nitric acid for a minimum of fifteen minutes, rinsed with reagent water, and dried in a clean environment
3. Sample Digestion
  - a. Tare the PFA digestion vessel.
  - b. Weigh out 0.1 g paint chip sample to the nearest 0.001 g into the tared PFA sample vessel. With large paint chip samples, measure out a 2 cm<sup>2</sup> piece, weigh to the nearest 0.001 g, and quantitatively transfer it to the vessel.

- c. Add  $5.0 \pm 0.1$  mL concentrated nitric acid to the sample vessel in a fume hood. If a vigorous reaction occurs, allow the reaction to stop before capping the vessel. Cap the vessel and torque the cap to 16 N-m according to the manufacturer's directions. The sample vessel may be connected to an overflow vessel using PFA connecting tubes. Place the vessels in the microwave carousel. Connect the overflow vessels to the center well of the unit.
- d. Place the vessels evenly distributed in the turntable of the microwave unit. Any vessels containing 5 mL of nitric acid for reagent blank purposes are counted as sample vessels. When fewer than the recommended number of samples are to be digested, i.e., three samples plus one blank, the remaining vessels should be filled with 5 mL of nitric acid to achieve the full complement of vessels. This provides an energy balance since the microwave power absorbed is proportional to the total mass in the cavity [17]. Irradiate each group of samples to achieve a temperature of 180 °C. Temperature ramp times should be appropriate for the vessels used. A sample digestion program for 12 samples is presented in Table A1. Adjust power values depending upon the number of samples included in the microwave at one time.

TABLE A1. PROGRAM VARIABLES FOR PAINT CHIPS SAMPLE DIGESTION WITH NITRIC ACID<sup>A</sup>

	<b>Stage 1</b>	<b>Stage 2</b>	<b>Stage 3</b>
<b>Power</b>	90%	90%	0%
<b>Pressure (kPa)</b>	375	750	---
<b>Run time (min)</b>	10	20	5
<b>Time @ P (min)</b>	5	15	---
<b>Temperature (°C)</b>	180	180	None applied

<sup>A</sup>For 12 microwave vessels that contain 0.1 g of sample and 5 mL of liquid per vessel

- e. At the end of the microwave program, allow the vessels to cool to a temperature below the boiling point of concentrated nitric acid (or that of the acid mixture used) before removing them from the microwave unit. If sample loss is detected (e.g., material in overflow collection vessel, liquid outside liner), determine the reason for the loss (e.g., loss of vessel seal integrity, use of a digestion time longer than 30 minutes, too large a sample, or improper heating conditions). Once the source of the loss has been corrected, prepare a new sample beginning at step 2. If insufficient material is available for reanalysis, dilute remaining digestate and note that some sample loss may have occurred.
  - f. Uncap and vent each vessel in a fume hood. Transfer the sample to an acid-cleaned polyethylene bottle. Dilute to 25 mL using reagent water. If the digested sample contains particulates which may clog nebulizers or interfere with injection of the sample into the instrument, allow the sample to settle or filter it:
    - Settling: Allow the sample to stand until the supernatant is clear (usually overnight is sufficient). If it does not filter the sample.
    - Filtering: Filter using disposable syringe filters, filter apparatus, etc. The filtering apparatus must be thoroughly precleaned and rinsed with dilute nitric acid. Filter the sample through quantitative filter paper into a second acid-cleaned container.
 The digestate is now ready for analysis for elements of interest using the appropriate method.
4. Calculations: Report the concentrations based on the actual weight of the original sample.